

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5374 SAS No.: P5374 SDG No.: P5374

Instrument ID: MSVOA_X Calibration Date/Time: 12/23/2024 08:49

Lab File ID: VX044466.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.739	0.733	0.1	-0.81	20
Vinyl Chloride	0.770	0.718		-6.75	20
Bromomethane	0.546	0.480		-12.09	20
Chloroethane	0.480	0.398		-17.08	20
Trichlorofluoromethane	1.508	1.293		-14.26	20
1,1,2-Trichlorotrifluoroethane	0.610	0.663		8.69	20
1,1-Dichloroethene	0.560	0.569		1.61	20
Acetone	0.357	0.295		-17.37	20
Carbon Disulfide	1.065	0.991		-6.95	20
Methyl tert-butyl Ether	2.187	1.927		-11.89	20
Methylene Chloride	0.671	0.595		-11.33	20
trans-1,2-Dichloroethene	0.596	0.569		-4.53	20
1,1-Dichloroethane	1.172	1.079	0.1	-7.93	20
2-Butanone	0.508	0.426		-16.14	20
Carbon Tetrachloride	0.482	0.532		10.37	20
cis-1,2-Dichloroethene	0.756	0.701		-7.28	20
Chloroform	1.316	1.165		-11.47	20
1,1,1-Trichloroethane	1.093	1.033		-5.49	20
Methylcyclohexane	0.535	0.657		22.8	20
Benzene	1.359	1.479		8.83	20
1,2-Dichloroethane	0.560	0.539		-3.75	20
Trichloroethene	0.339	0.375		10.62	20
1,2-Dichloropropane	0.347	0.363		4.61	20
Bromodichloromethane	0.465	0.505		8.6	20
4-Methyl-2-Pentanone	0.556	0.519		-6.66	20
Toluene	0.853	0.904		5.98	20
t-1,3-Dichloropropene	0.455	0.481		5.71	20
cis-1,3-Dichloropropene	0.507	0.551		8.68	20
1,1,2-Trichloroethane	0.339	0.361		6.49	20
2-Hexanone	0.403	0.376		-6.7	20
Dibromochloromethane	0.326	0.354		8.59	20
Tetrachloroethene	0.332	0.399		20.18	20
Chlorobenzene	1.071	1.145	0.3	6.91	20
Ethyl Benzene	1.851	1.931		4.32	20
m/p-Xylenes	0.679	0.747		10.02	20
o-Xylene	0.687	0.732		6.55	20
Styrene	1.102	1.199		8.8	20
Bromoform	0.219	0.252	0.1	15.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5374 SAS No.: P5374 SDG No.: P5374

Instrument ID: MSVOA_X Calibration Date/Time: 12/23/2024 08:49

Lab File ID: VX044466.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.828	4.164		8.78	20
1,1,2,2-Tetrachloroethane	1.329	1.384	0.3	4.14	20
1,3-Dichlorobenzene	1.621	1.739		7.28	20
1,4-Dichlorobenzene	1.667	1.793		7.56	20
1,2-Dichlorobenzene	1.684	1.789		6.24	20
1,2-Dichloroethane-d4	0.877	0.735		-16.19	20
Dibromofluoromethane	0.346	0.372		7.51	20
Toluene-d8	1.168	1.249		6.93	20
4-Bromofluorobenzene	0.408	0.415		1.72	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.