

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P5374 SAS No.: P5374 SDG NO.: P5374
Lab File ID: VX044466.D Date Analyzed: 12/23/2024
Instrument ID: MSVOA_X Time Analyzed: 08:49
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98270	5.54	155098	6.75	129773	10.05
UPPER LIMIT	196540	6.044	310196	7.251	259546	10.549
LOWER LIMIT	49135	5.044	77549	6.251	64886.5	9.549
EPA SAMPLE NO.						
BP-VPB-190A-TB-20241218	144905	5.54	250665	6.76	212481	10.06
BP-VPB-190A-EB-20241218	136476	5.54	243949	6.76	221103	10.05
BP-VPB-190A-GW-718-720	132615	5.54	236368	6.76	212205	10.05
BP-VPB-190A-GW-748-750	132243	5.54	237676	6.76	212650	10.05
BP-VPB-190A-GW-748-750MS	131546	5.54	214972	6.76	185795	10.05
BP-VPB-190A-GW-748-750MSD	133309	5.55	221334	6.76	194386	10.05
VX1223WBL01	153955	5.54	275095	6.76	235868	10.05
VX1223WBS03	136426	5.54	230675	6.76	197773	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.