

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P5374 SAS No.: P5374 SDG NO.: P5374
Lab File ID: VX044466.D Date Analyzed: 12/23/2024
Instrument ID: MSVOA_X Time Analyzed: 08:49
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	59280	12.018				
UPPER LIMIT	118560	12.518				
LOWER LIMIT	29640	11.518				
EPA SAMPLE NO.						
BP-VPB-190A-TB-20241218	90686	12.02				
BP-VPB-190A-EB-20241218	90813	12.02				
BP-VPB-190A-GW-718-720	92424	12.02				
BP-VPB-190A-GW-748-750	92633	12.02				
BP-VPB-190A-GW-748-750MS	87327	12.02				
BP-VPB-190A-GW-748-750MSD	93024	12.02				
VX1223WBL01	97837	12.02				
VX1223WBS03	93712	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.