


LB134084

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.077	0.077	-145	0.000	-145	0				0/00
0.2	0.200	0.216	0.016	447	0.0 %	447					8
2.5	2.500	2.466	-0.034	10028	0.0 %	10028					-1
5.0	5.000	4.936	-0.064	20541	0.0 %	20541					-1
7.5	7.500	7.377	-0.123	30936	0.0 %	30936					-2
10.0	10.000	10.128	0.128	42646	0.0 %	42646					1

LB134084 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Dec 2024 10:28:06

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	-145	-	0.0000	7470A	PPB	26 Dec 2024 10:38:25	S	Std
0.2 - 1	50.2	447	-	0.2000	7470A	PPB	26 Dec 2024 10:40:43	S	Std
2.5 - 1	52.5	10028	-	2.5000	7470A	PPB	26 Dec 2024 10:43:00	S	Std
5.0 - 1	55	20541	-	5.0000	7470A	PPB	26 Dec 2024 10:45:17	S	Std
7.5 - 1	57.5	30936	-	7.5000	7470A	PPB	26 Dec 2024 10:47:35	S	Std
10.0 - 1	516	42646	-	10.0000	7470A	PPB	26 Dec 2024 10:49:53	S	Std
ICV30 - 1	ICV30	15768	3.8146	-	7470A	PPB	26 Dec 2024 10:52:50	U	SMPL
ICB30 - 1	ICB30	-216	0.0601	-	7470A	PPB	26 Dec 2024 10:55:09	U	SMPL
CCV64 - 1	CCV64	20615	4.9531	-	7470A	PPB	26 Dec 2024 10:57:21	U	SMPL
CCB64 - 1	CCB64	-213	0.0608	-	7470A	PPB	26 Dec 2024 10:59:36	U	SMPL
CRA - 1	CRA	503	0.2290	-	7470A	PPB	26 Dec 2024 11:07:40	U	SMPL
HighStd - 1	HighStd	42378	10.0650	-	7470A	PPB	26 Dec 2024 11:09:55	U	SMPL
ChkStd - 1	ChkStd	28294	6.7568	-	7470A	PPB	26 Dec 2024 11:12:10	U	SMPL
PB165856BL - 1	PBW	-309	0.0383	-	7470A	PPB	26 Dec 2024 11:14:28	U	SMPL
PB165856BS - 1	LCSW	15636	3.7836	-	7470A	PPB	26 Dec 2024 11:19:22	U	SMPL
P5375-01 - 1	TT-RW10A-IDWGW-20241219	-172	0.0705	-	7470A	PPB	26 Dec 2024 11:21:38	U	SMPL
P5375-01DUP - 1	TT-RW10A-IDWGW-20241219DUP	-24	0.1052	-	7470A	PPB	26 Dec 2024 11:23:57	U	SMPL
P5375-01MS - 1	TT-RW10A-IDWGW-20241219MS	15593	3.7735	-	7470A	PPB	26 Dec 2024 11:28:52	U	SMPL
P5375-01MSD - 1	TT-RW10A-IDWGW-20241219MSD	15677	3.7932	-	7470A	PPB	26 Dec 2024 11:31:08	U	SMPL
CCV65 - 1	CCV65	20571	4.9428	-	7470A	PPB	26 Dec 2024 11:33:25	U	SMPL
CCB65 - 1	CCB65	-242	0.0540	-	7470A	PPB	26 Dec 2024 11:35:42	U	SMPL
P5375-01LX5 - 1		29	0.1177	-	7470A	PPB	26 Dec 2024 11:37:59	U	SMPL
P5375-01A - 1		15653	3.7876	-	7470A	PPB	26 Dec 2024 11:42:54	U	SMPL
CCV66 - 1	CCV66	20284	4.8753	-	7470A	PPB	26 Dec 2024 11:45:10	U	SMPL
CCB66 - 1	CCB66	-239	0.0547	-	7470A	PPB	26 Dec 2024 11:47:26	U	SMPL