

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P5375

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165856BL		WATER		Batch Number:		PB165856		Prep Date:	12/26/2024	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	12/26/2024	11:14	LB134084

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P5375

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB165832BL	WATER			Batch Number:		PB165832		Prep Date:	12/23/2024	
	Aluminum	50.0	<50.0	U	40.0	50.0	P	12/27/2024	16:20	LB134113
	Antimony	25.0	<25.0	U	6.25	25.0	P	12/27/2024	16:20	LB134113
	Arsenic	10.0	<10.0	U	8.00	10.0	P	12/27/2024	16:20	LB134113
	Barium	50.0	<50.0	U	12.5	50.0	P	12/27/2024	16:20	LB134113
	Beryllium	3.00	<3.00	U	0.75	3.00	P	12/27/2024	16:20	LB134113
	Cadmium	3.00	<3.00	U	0.75	3.00	P	12/27/2024	16:20	LB134113
	Calcium	1000	<1000	U	250	1000	P	12/27/2024	16:20	LB134113
	Chromium	5.00	<5.00	U	2.50	5.00	P	12/27/2024	16:20	LB134113
	Cobalt	15.0	<15.0	U	3.75	15.0	P	12/27/2024	16:20	LB134113
	Copper	10.0	<10.0	U	8.00	10.0	P	12/27/2024	16:20	LB134113
	Iron	50.0	<50.0	U	40.0	50.0	P	12/27/2024	16:20	LB134113
	Lead	6.00	<6.00	U	4.80	6.00	P	12/27/2024	16:20	LB134113
	Magnesium	1000	<1000	U	250	1000	P	12/27/2024	16:20	LB134113
	Manganese	10.0	<10.0	U	2.50	10.0	P	12/27/2024	16:20	LB134113
	Nickel	20.0	<20.0	U	5.00	20.0	P	12/27/2024	16:20	LB134113
	Potassium	1000	<1000	U	800	1000	P	12/27/2024	16:20	LB134113
	Selenium	10.0	<10.0	U	8.00	10.0	P	12/27/2024	16:20	LB134113
	Silver	5.00	<5.00	U	2.50	5.00	P	12/27/2024	16:20	LB134113
	Sodium	1000	<1000	U	500	1000	P	12/27/2024	16:20	LB134113
	Thallium	20.0	<20.0	U	10.0	20.0	P	12/27/2024	16:20	LB134113
	Vanadium	20.0	<20.0	U	10.0	20.0	P	12/27/2024	16:20	LB134113
	Zinc	20.0	<20.0	U	5.00	20.0	P	12/27/2024	16:20	LB134113