

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P5375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5375 **SAS No.:** P5375
Matrix: Water **Sample ID:** P5375-01 **Client ID:** TT-RW10A-IDWGW-20241219DUP
Percent Solids for Sample: NA **Duplicate ID** P5375-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1470		1530		4		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	14.4	J	14.5	J	1		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	0.13	J	0.097	J	31		P
Calcium	ug/L	20	7580		7550		0		P
Chromium	ug/L	20	27.3		27.2		0		P
Cobalt	ug/L	20	19.7		19.7		0		P
Copper	ug/L	20	39.5		39.7		1		P
Iron	ug/L	20	2450		2440		0		P
Lead	ug/L	20	6.00	U	6.00	U			P
Magnesium	ug/L	20	3480		3450		1		P
Manganese	ug/L	20	109		109		0		P
Mercury	ug/L	20	0.20	U	0.11	J	200.0		CV
Nickel	ug/L	20	37.5		37.7		1		P
Potassium	ug/L	20	2940		3010		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	1.87	J	1.86	J	1		P
Sodium	ug/L	20	50300		51700		3		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	79.1		78.8		0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P5375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5375 **SAS No.:** P5375
Matrix: Water **Sample ID:** P5375-01MS **Client ID:** TT-RW10A-IDWGW-20241219MSD
Percent Solids for Sample: NA **Duplicate ID** P5375-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	2490		2480		0		P
Antimony	ug/L	20	426		429		1		P
Arsenic	ug/L	20	422		424		0		P
Barium	ug/L	20	133		131		2		P
Beryllium	ug/L	20	77.7		76.9		1		P
Cadmium	ug/L	20	81.5		81.5		0		P
Calcium	ug/L	20	7800		7710		1		P
Chromium	ug/L	20	188		188		0		P
Cobalt	ug/L	20	110		110		0		P
Copper	ug/L	20	187		188		1		P
Iron	ug/L	20	3700		3700		0		P
Lead	ug/L	20	432		432		0		P
Magnesium	ug/L	20	4280		4250		1		P
Manganese	ug/L	20	188		187		1		P
Mercury	ug/L	20	3.77		3.79		1		CV
Nickel	ug/L	20	250		250		0		P
Potassium	ug/L	20	8490		8440		1		P
Selenium	ug/L	20	1020		1030		1		P
Silver	ug/L	20	36.6		36.3		1		P
Sodium	ug/L	20	52300		52500		0		P
Thallium	ug/L	20	950		928		2		P
Vanadium	ug/L	20	136		135		1		P
Zinc	ug/L	20	179		179		0		P