

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** P5375
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P5375 **SAS No.:** P5375

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165832BS							
Aluminum	ug/L	1000	939		94	86 - 115	P
Antimony	ug/L	400	406		102	88 - 113	P
Arsenic	ug/L	400	414		104	87 - 113	P
Barium	ug/L	100	89.1		89	88 - 113	P
Beryllium	ug/L	100	91.8		92	89 - 112	P
Cadmium	ug/L	100	95.6		96	88 - 113	P
Calcium	ug/L	500	478	J	96	87 - 113	P
Chromium	ug/L	200	202		101	90 - 113	P
Cobalt	ug/L	100	98.0		98	89 - 114	P
Copper	ug/L	150	160		107	86 - 114	P
Iron	ug/L	1500	1460		97	87 - 115	P
Lead	ug/L	500	487		97	86 - 113	P
Magnesium	ug/L	1000	905	J	90	85 - 113	P
Manganese	ug/L	100	91.0		91	90 - 114	P
Nickel	ug/L	250	245		98	88 - 113	P
Potassium	ug/L	5000	4860		97	86 - 114	P
Selenium	ug/L	1000	1020		102	83 - 114	P
Silver	ug/L	37.5	38.4		102	84 - 115	P
Sodium	ug/L	1500	1420		95	87 - 115	P
Thallium	ug/L	1000	1020		102	85 - 114	P
Vanadium	ug/L	150	140		93	90 - 111	P
Zinc	ug/L	100	94.8		95	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P5375

Contract: TETR06

Lab Code: CHEM

Case No.: P5375

SAS No.: P5375

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165856BS Mercury	ug/L	4.0	3.78		94	82 - 119	CV