

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

TT-RW10A-IDWGW-20241219L

Lab Name: Chemtech Consulting Group **Contract:** TETR06
Lab Code: CHEM **Lb No.:** lb134090 **Lab Sample ID :** P5375-01L **SDG No.:** P5375
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Aluminum	1470		1530		4		P
Antimony	25.0	U	125	U			P
Arsenic	10.0	U	50.0	U			P
Barium	14.4	J	250	U	100.0		P
Beryllium	3.00	U	15.0	U			P
Cadmium	0.13	J	15.0	U	100.0		P
Calcium	7580		7800		3		P
Chromium	27.3		27.5		1		P
Cobalt	19.7		20.3	J	3		P
Copper	39.5		42.6	J	8		P
Iron	2450		2440		0		P
Lead	6.00	U	30.0	U			P
Magnesium	3480		3570	J	3		P
Manganese	109		112		2		P
Mercury	0.20	U	0.59	J	100.0		CV
Nickel	37.5		38.7	J	3		P
Potassium	2940		5000	U	100.0		P
Selenium	10.0	U	50.0	U			P
Silver	1.87	J	9.74	J	422		P
Sodium	50300		51300		2		P
Thallium	20.0	U	100	U			P
Vanadium	20.0	U	100	U			P
Zinc	79.1		77.9	J	1		P