

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/19/24
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/20/24
Client Sample ID:	TT162S1-20241219		SDG No.:	P5376
Lab Sample ID:	P5376-15		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035847.D	1	12/23/24 13:29	12/26/24 11:20	PB165828

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.34		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		98%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.46	*	55 - 111		114%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		98%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		134%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	3700	7.257				
1146-65-2	Naphthalene-d8	8880	9.999				
15067-26-2	Acenaphthene-d10	5150	13.914				
1517-22-2	Phenanthrene-d10	10600	16.686				
1719-03-5	Chrysene-d12	8140	20.947				
1520-96-3	Perylene-d12	8010	23.026				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products