

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5376 SAS No.: P5376 SDG NO.: P5376

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/24/2024

Lab File ID: BN035830.D Time Analyzed: 20:55

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3866	7.257	9955	10.00	5956	13.91
UPPER LIMIT	7732	7.757	19910	10.498	11912	14.414
LOWER LIMIT	1933	6.757	4977.5	9.498	2978	13.414
EPA SAMPLE NO.						
01 RW11-MW01S-20241219	3277	7.26	7724	10.01	4815	13.91
02 RW11-MW01I-20241219	3211	7.26	7680	10.01	4788	13.93
03 TT119S1-20241219	3270	7.26	7393	10.01	4622	13.93
04 FB03-20241219	2951	7.26	6785	10.01	4279	13.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.