

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/18/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/20/24	
Client Sample ID:	RW10-MW01S-20241218MS		SDG No.:	P5376	
Lab Sample ID:	P5376-08MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044564.D	1		12/30/24 23:08	VX123124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	46.4		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	46.4		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	45.2		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	42.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	41.3		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	42.3		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	47.8		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	51.2		0.26	0.75	1.00	ug/L
67-64-1	Acetone	220		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	43.3		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	46.7		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	45.7		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	49.7		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	49.3		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	49.2		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	45.4		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	240		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	51.6		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	48.0		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	42.6		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	46.7		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	47.4		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	49.9		0.19	0.50	1.00	ug/L
71-43-2	Benzene	52.7		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	50.2		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	53.6		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	50.4		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	53.9		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	260		0.75	2.50	5.00	ug/L
108-88-3	Toluene	52.3		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	12/18/24
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	12/20/24
Client Sample ID:	RW10-MW01S-20241218MS	SDG No.:	P5376
Lab Sample ID:	P5376-08MS	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044564.D	1		12/30/24 23:08	VX123124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	47.2		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	50.4		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	54.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	270		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	51.8		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	54.5		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	51.8		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	51.1		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	49.6		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	100		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	50.3		0.14	0.50	1.00	ug/L
100-42-5	Styrene	53.3		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	51.2		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	49.2		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	49.7		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	51.2		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	50.3		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	51.9		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	49.4		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	53.7		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	52.7		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	41.3		81 - 118		83%	SPK: 50
1868-53-7	Dibromofluoromethane	48.1		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	47.9		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.9		85 - 114		98%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	121000	5.55				
540-36-3	1,4-Difluorobenzene	205000	6.757				
3114-55-4	Chlorobenzene-d5	186000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	88800	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/18/24	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/20/24	
Client Sample ID:	RW10-MW01S-20241218MS		SDG No.:	P5376	
Lab Sample ID:	P5376-08MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044564.D	1		12/30/24 23:08	VX123124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products