



Lab Name:	<u>CHEMTECH</u>			Contract:	<u>AECO15</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5376</u>	SAS No.:	<u>P5376</u>	SDG NO.:	<u>P5376</u>
Lab File ID:	<u>VX044540.D</u>			Date Analyzed:	<u>12/30/2024</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>13:07</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137969	5.54	223553	6.76	195826	10.06
UPPER LIMIT	275938	6.044	447106	7.257	391652	10.555
LOWER LIMIT	68984.5	5.044	111777	6.257	97913	9.555
EPA SAMPLE NO.						
TT205S1-20241218	125750	5.55	210247	6.76	190422	10.06
TT150S1-20241218	123263	5.55	213060	6.76	182287	10.06
RE115D2-20241218	118700	5.55	207183	6.76	174651	10.06
TT149S1-20241218	121289	5.55	209255	6.76	185772	10.06
TT174I1-20241218	127903	5.55	216454	6.76	194600	10.05
RW10-MW01S-20241218	109256	5.55	192338	6.76	172631	10.06
RW10-MW01S-20241218MS	121315	5.55	205358	6.76	185869	10.06
RW10-MW01S-20241218MSD	117366	5.55	203568	6.76	185270	10.06
RW10-MW01D-20241218	123900	5.55	212555	6.76	188273	10.06
RW10A-MW01S-20241218	122970	5.55	208916	6.76	183505	10.06
RW10A-MW01I-20241218	119008	5.54	207504	6.76	180793	10.05
RE117D2-20241218	114236	5.54	200838	6.76	178297	10.06
TT161S1-20241219	105027	5.55	186168	6.76	163359	10.06
TT162S1-20241219	114172	5.55	199344	6.76	174896	10.06
RW11-MW01S-20241219	109001	5.54	187510	6.76	162626	10.06
RW11-MW01I-20241219	114745	5.55	197414	6.76	174241	10.05
TT119S1-20241219	119469	5.55	211589	6.76	185760	10.06
TB	117791	5.55	206326	6.76	188782	10.05
VX1231WBL01	117464	5.55	200538	6.76	172503	10.06
VX1231WBS01	153746	5.54	254154	6.76	219785	10.06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5376 SAS No.: P5376 SDG NO.: P5376
 Lab File ID: VX044540.D Date Analyzed: 12/30/2024
 Instrument ID: MSVOA_X Time Analyzed: 13:07
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137969	5.54	223553	6.76	195826	10.06
UPPER LIMIT	275938	6.044	447106	7.257	391652	10.555
LOWER LIMIT	68984.5	5.044	111777	6.257	97913	9.555
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.