

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5376 SAS No.: P5376 SDG NO.: P5376
Lab File ID: VX044568.D Date Analyzed: 12/31/2024
Instrument ID: MSVOA_X Time Analyzed: 10:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	126884	5.54	200592	6.76	176322	10.05
UPPER LIMIT	253768	6.038	401184	7.257	352644	10.549
LOWER LIMIT	63442	5.038	100296	6.257	88161	9.549
EPA SAMPLE NO.						
DUP-07-20241218	126196	5.54	215542	6.76	190830	10.06
TT188S1-20241219	131976	5.55	226699	6.76	193617	10.06
FB03-20241219	125076	5.54	211805	6.76	185221	10.06
VX1231WBL02	128703	5.54	217401	6.76	189142	10.05
VX1231WBS02	157790	5.54	262850	6.76	220076	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.