

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>			Contract:	<u>AECO15</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5376</u>	SAS No.:	<u>P5376</u>	SDG NO.:	<u>P5376</u>
Lab File ID:	<u>VX044540.D</u>			Date Analyzed:	<u>12/30/2024</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>13:07</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS4 AREA #	RT #				
12 HOUR STD	98470	12.018				
UPPER LIMIT	196940	12.518				
LOWER LIMIT	49235	11.518				
EPA SAMPLE NO.						
TT205S1-20241218	80361	12.02				
TT150S1-20241218	76648	12.02				
RE115D2-20241218	70024	12.02				
TT149S1-20241218	82258	12.02				
TT174I1-20241218	84267	12.02				
RW10-MW01S-20241218	69122	12.02				
RW10-MW01S-20241218MS	88817	12.02				
RW10-MW01S-20241218MSD	88862	12.02				
RW10-MW01D-20241218	80341	12.02				
RW10A-MW01S-20241218	75781	12.02				
RW10A-MW01I-20241218	73748	12.02				
RE117D2-20241218	70367	12.02				
TT161S1-20241219	66627	12.02				
TT162S1-20241219	74318	12.02				
RW11-MW01S-20241219	63684	12.02				
RW11-MW01I-20241219	75591	12.02				
TT119S1-20241219	75337	12.02				
TB	77501	12.02				
VX1231WBL01	71731	12.02				
VX1231WBS01	100843	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5376 SAS No.: P5376 SDG NO.: P5376
 Lab File ID: VX044540.D Date Analyzed: 12/30/2024
 Instrument ID: MSVOA_X Time Analyzed: 13:07
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	98470	12.018				
UPPER LIMIT	196940	12.518				
LOWER LIMIT	49235	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.