

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5376 SAS No.: P5376 SDG NO.: P5376
 Lab File ID: VX044568.D Date Analyzed: 12/31/2024
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	86636	12.018				
UPPER LIMIT	173272	12.518				
LOWER LIMIT	43318	11.518				
EPA SAMPLE NO.						
DUP-07-20241218	83236	12.02				
TT188S1-20241219	80709	12.02				
FB03-20241219	85230	12.02				
VX1231WBL02	80019	12.02				
VX1231WBS02	100083	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.