

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134101

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/27/2024	09:37
2	CAL2	1	Water	NA	NA	20.2	7.01	12/27/2024	09:38
3	CAL3	1	Water	NA	NA	20.3	10.02	12/27/2024	09:40
4	ICV	1	Water	NA	NA	20.3	7.00	12/27/2024	09:44
5	CCV1	1	Water	NA	NA	20.2	2.01	12/27/2024	09:45
6	P5380-01	1	Solid	20.02	20	20.2	10.54	12/27/2024	09:55
7	P5380-01DUP	1	Solid	20.03	20	20.3	10.55	12/27/2024	09:56
8	CCV2	1	Water	NA	NA	20.3	12.02	12/27/2024	10:00

WORKLIST(Hardcopy Internal Chain)

97134101

WorkList Name : ph p5380 WorkList ID : 186659 Department : Wet-Chemistry Date : 12-27-2024 09:24:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Solid	pH	Cool 4 deg C	WEST04	N31	12/20/2024	9045D

Date/Time 12/27/24 09:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 12/27/24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]