

## **DATA PACKAGE GENERAL CHEMISTRY**

**PROJECT NAME : FT MEADE TIPTON AIRFIELD PARCEL RI - PO 0111169**

**WESTON SOLUTIONS**

**1400 Weston Way**

**PO Box 2653**

**West Chester, PA - 19380**

**Phone No: 610-701-7400**

**ORDER ID : P5380**

**ATTENTION : Nathan Fretz**



**Laboratory Certification ID # 20012**



1) GENERAL CHEMISTRY DATA	2
2) Signature Page	3
3) Case Narrative	4
4) Qualifier Page	6
5) Conformance/Non Conformance	7
6) QA Checklist	8
7) Chronicle	9
8) Sample Data	10
8.1) TAPIAL3-IDW-SOIL-122024-T1	11
9) QC Data Summary For Genchem	12
9.1) Initial and Continuing Calibration Verification	13
9.2) Initial and Continuing Calibration Blank Summary	15
9.3) Preparation Blank Summary	16
9.4) Matrix Spike Summary	17
9.5) Duplicate Sample Summary	19
9.6) Laboratory Control Sample Summary	23
10) GENCHEM RAW DATA	25
10.1) GENCHEM RAW DATA - ANALYTICAL	26
10.1.1) LB134101	26
10.1.2) LB134107	28
10.1.3) LB134121	30
10.1.4) LB134137	31
10.2) GENCHEM RAW DATA - PREP	34
10.2.1) PB165915	34
10.2.2) PB165932	37
11) Analytical Runlogs	40
12) Standard Prep Logs	45
13) Percent Solid	97
14) Shipping Document	100
14.1) Chain Of Custody	101
14.2) Lab Certificate	102

## Cover Page

**Order ID :** P5380

**Project ID :** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**Client :** Weston Solutions

**Lab Sample Number**

P5380-01  
P5380-02

**Client Sample Number**

TAPIAL3-IDW-SOIL-122024-T1  
TAPIAL3-IDW-SOIL-122024-T1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 1/9/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Weston Solutions**

**Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169**

**Project # N/A**

**Chemtech Project # P5380**

**Test Name: pH,Cyanide,Sulfide,Ignitability**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 12/21/2024.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Cyanide, Ignitability, PCB, pH, Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP Pesticide, TCLP VOA and TCLP ZHE Extraction. This data package contains results for pH, Cyanide, Sulfide, Ignitability.

### **C. Analytical Techniques:**

The analysis of Ignitability was based on method 1030, The analysis of Cyanide was based on method 9012B, The analysis of Sulfide was based on method 9034 and The analysis of pH was based on method 9045D.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for TAPIAL3-IDW-SOIL-122024-T1 of pH as sample receive out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

#### **Calculation for CN Soil Sample:**

Conversion of Results from  $\mu\text{g} / \text{L}$  or  $\text{ppb}$  to  $\text{mg}/\text{kg}$ :

$$\text{Concentration (mg/kg)} = C \times \frac{V_f}{W \times S} \times \text{DF} / 1000$$





Where,

C = Instrument response in  $\mu\text{g/L}$  CN from the calibration curve.

Vf = Final prepared (absorbing solution) volume (mL)

W = Initial aliquot amount (g) (Fraction of Sample amount taken in prep)

S = % Solids / 100 (Fraction of Percent Solids)

DF = Dilution Factor

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature \_\_\_\_\_

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: P5380

MATRIX: Solid

METHOD: 1030,9012B,9034,9045D

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met		✓	
If not met, list number of days exceeded for each sample:			
The Holding Times were met for all samples except for TAPIAL3-IDW-SOIL-122024-T1 of pH as sample receive out of holding time.			

ADDITIONAL COMMENTS:

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date

APPENDIX A

**QA REVIEW GENERAL DOCUMENTATION**

Project #: P5380

Completed

For thorough review, the report must have the following:

**GENERAL:**

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

**COVER PAGE:**

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

**CHAIN OF CUSTODY:**

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

**ANALYTICAL:**

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 01/09/2025

LAB CHRONICLE

OrderID:	P5380	OrderDate:	12/23/2024 9:50:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5380-01	TAPIAL3-IDW-SOIL-1 22024-T1	SOIL			12/20/24 14:15			12/21/24
			Cyanide	9012B		12/31/24	12/31/24 12:38	
			Ignitability	1030			12/27/24 08:30	
			pH	9045D			12/27/24 09:55	
			Sulfide	9034		12/30/24	12/30/24 13:26	



# SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14

## Report of Analysis

Client:	Weston Solutions	Date Collected:	12/20/24 14:15
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	12/21/24
Client Sample ID:	TAPIAL3-IDW-SOIL-122024-T1	SDG No.:	P5380
Lab Sample ID:	P5380-01	Matrix:	SOIL
		% Solid:	86.3

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Cyanide	0.23	U	1	0.049	0.23	0.28	mg/Kg	12/31/24 08:50	12/31/24 12:38	9012B
Ignitability	NO		1	0	0	0	oC		12/27/24 08:30	1030
pH	10.5	H	1	0	0	0	pH		12/27/24 09:55	9045D
Sulfide	3.70	J	1	2.15	5.77	11.5	mg/Kg	12/30/24 08:45	12/30/24 13:26	9034

Comments: pH result reported at temperature 20.2 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



# QC RESULT SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14



## Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5380

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB134101

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV</b> pH	pH	7.00	7	100	90-110	12/27/2024
Sample ID: <b>CCV1</b> pH	pH	2.01	2.00	101	90-110	12/27/2024
Sample ID: <b>CCV2</b> pH	pH	12.02	12.00	100	90-110	12/27/2024

## Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5380

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB134137

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV1</b> Cyanide	mg/L	0.097	0.099	98	90-110	12/31/2024
Sample ID: <b>CCV1</b> Cyanide	mg/L	0.24	0.25	96	90-110	12/31/2024
Sample ID: <b>CCV2</b> Cyanide	mg/L	0.24	0.25	96	90-110	12/31/2024

### Initial and Continuing Calibration Blank Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>RunNo.:</b>	LB134137

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>ICB1</b> Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/31/2024
Sample ID: <b>CCB1</b> Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/31/2024
Sample ID: <b>CCB2</b> Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/31/2024

### Preparation Blank Summary

**Client:** Weston Solutions

**SDG No.:** P5380

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: Sulfide	<b>PB165915BL</b> mg/Kg	< 5.0000	5.0000	U	1.86	10.0	12/30/2024
Sample ID: Cyanide	<b>PB165932BL</b> mg/Kg	< 0.1250	0.1250	U	0.044	0.25	12/31/2024

Matrix Spike Summary

Client:

Weston Solutions

SDG No.:

P5380

Project:

Ft Meade Tipton Airfield Parcel RI - PO 0111169

Sample ID:

P5380-01

Client ID:

TAPIAL3-IDW-SOIL-122024-T1MS

Percent Solids for Spike Sample:

86.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Sulfide	mg/Kg	75-125	236		3.70	J	290	1	80		12/30/2024
Cyanide	mg/Kg	75-125	1.80		0.049	U	2.3	1	78		12/31/2024

## Matrix Spike Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5380-01
<b>Client ID:</b>	TAPIAL3-IDW-SOIL-122024-T1MSD	<b>Percent Solids for Spike Sample:</b>	86.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Sulfide	mg/Kg	75-125	234		3.70	J	290	1	79		12/30/2024
Cyanide	mg/Kg	75-125	1.80		0.049	U	2.3	1	78		12/31/2024

Duplicate Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5380-01
<b>Client ID:</b>	TAPIAL3-IDW-SOIL-122024-T1DUP	<b>Percent Solids for Spike Sample:</b>	86.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
pH	pH	+/-20	10.5		10.6		1	0.09		12/27/2024

Duplicate Sample Summary

**Client:** Weston Solutions

**SDG No.:** P5380

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**Sample ID:** P5380-01

**Client ID:** TAPIAL3-IDW-SOIL-122024-T1DUP

**Percent Solids for Spike Sample:** 86.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Sulfide	mg/Kg	+/-20	3.70	J	3.70	J	1	0		12/30/2024
Cyanide	mg/Kg	+/-20	0.049	U	0.049	U	1	0		12/31/2024



Duplicate Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5380-01
<b>Client ID:</b>	TAPIAL3-IDW-SOIL-122024-T1MSD	<b>Percent Solids for Spike Sample:</b>	86.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Sulfide	mg/Kg	+/-20	236		234		1	0.85		12/30/2024
Cyanide	mg/Kg	+/-20	1.80		1.80		1	0		12/31/2024

### Duplicate Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5386-04
<b>Client ID:</b>	MOO-24-00395-96DUP	<b>Percent Solids for Spike Sample:</b>	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		12/27/2024

### Laboratory Control Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Run No.:</b>	LB134121

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165915BS							
Sulfide	mg/Kg	250	219		88	1	80-120	12/30/2024

### Laboratory Control Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5380
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Run No.:</b>	LB134137

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165932BS							
Cyanide	mg/Kg	5	4.70		94	1	85-115	12/31/2024



# RAW DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134101

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/27/2024	09:37
2	CAL2	1	Water	NA	NA	20.2	7.01	12/27/2024	09:38
3	CAL3	1	Water	NA	NA	20.3	10.02	12/27/2024	09:40
4	ICV	1	Water	NA	NA	20.3	7.00	12/27/2024	09:44
5	CCV1	1	Water	NA	NA	20.2	2.01	12/27/2024	09:45
6	P5380-01	1	Solid	20.02	20	20.2	10.54	12/27/2024	09:55
7	P5380-01DUP	1	Solid	20.03	20	20.3	10.55	12/27/2024	09:56
8	CCV2	1	Water	NA	NA	20.3	12.02	12/27/2024	10:00

WORKLIST(Hardcopy Internal Chain)

97134101

WorkList Name : ph p5380      WorkList ID : 186659      Department : Wet-Chemistry      Date : 12-27-2024 09:24:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Solid	pH	Cool 4 deg C	WEST04	N31	12/20/2024	9045D

Date/Time 12/27/24 09:30  
Raw Sample Received by: [Signature]  
Raw Sample Relinquished by: [Signature]

Date/Time 12/27/24 12:00  
Raw Sample Received by: [Signature]  
Raw Sample Relinquished by: [Signature]

**Analytical Summary Report**

Analysis Method: 1030  
Parameter: Ignitability  
Run Number: LB134107

Reviewed By: rubina  
Supervisor Review By: Iwona

Seq	LabID	ClientID	DF	matrix	Result Status	Burning Rate	Anal Date	Anal Time
1	P5380-01	TAPIAL3-IDW-SOIL-122	1	Solid	NO	0.00	12/27/2024	08:30
2	P5386-01	MOO-24-00398	1	Solid	NO	0.00	12/27/2024	08:38
3	P5386-02	MOO-24-00398	1	Solid	NO	0.00	12/27/2024	08:45
4	P5386-03	MOO-24-00395-96	1	Solid	NO	0.00	12/27/2024	08:52
5	P5386-04	MOO-24-00395-96	1	Solid	NO	0.00	12/27/2024	09:00
6	P5386-04DUP	MOO-24-00395-96DUP	1	Solid	NO	0.00	12/27/2024	09:07

Burning Rate =  $\frac{\text{Length (mm)}}{\text{Total Time (sec)}}$

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14



WORKLIST(Hardcopy Internal Chain)

66134107

WorkList Name : ign-12-26

WorkList ID : 186606

Department : Wet-Chemistry

Date : 12-26-2024 08:21:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Solid	Ignitability	Cool 4 deg C	WEST04	N31	12/20/2024	1030
P5386-01	MOO-24-00398	Solid	Ignitability	Cool 4 deg C	PSEG03	N31	12/26/2024	1030
P5386-02	MOO-24-00398	Solid	Ignitability	Cool 4 deg C	PSEG03	N31	12/26/2024	1030
P5386-03	MOO-24-00395-96	Solid	Ignitability	Cool 4 deg C	PSEG03	N31	12/26/2024	1030
P5386-04	MOO-24-00395-96	Solid	Ignitability	Cool 4 deg C	PSEG03	N31	12/26/2024	1030

Date/Time 12/27/2024 08:15  
Raw Sample Received by: RIT cws  
Raw Sample Relinquished by: JH WDC

Date/Time 12/27/2024 09:22  
Raw Sample Received by: JH WDC  
Raw Sample Relinquished by: RIT cws

Analysis Method: 9034  
Parameter: Sulfide  
Run Number: LB134121

ANALYST: rubina  
SUPERVISOR REVIEW BY: Iwona  
Constant: 16000  
Normality1: 0.025  
Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3114
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/L)	DF	Initial Weight (g)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB165915BL		1	5.00	50	5.00	0.00	4.94	4.94	0.06	0.00	0.00	12/30/2024	13:20
2	PB165915BS	250	1	5.00	50	5.00	0.00	2.20	2.20	2.80	2.74	219.20	12/30/2024	13:23
3	P5380-01		1	5.02	50	5.00	0.00	4.90	4.90	0.10	0.04	3.19	12/30/2024	13:26
4	P5380-01DUP		1	5.02	50	5.00	0.00	4.90	4.90	0.10	0.04	3.19	12/30/2024	13:29
5	P5380-01MS	250	1	5.03	50	5.00	0.00	2.38	2.38	2.62	2.56	203.58	12/30/2024	13:32
6	P5380-01MSD	250	1	5.03	50	5.00	0.00	2.40	2.40	2.60	2.54	201.99	12/30/2024	13:35

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume

6134137

Test results

Aquakem 7.2AQ1

Page:

CHEMTECH CONSULTING GROUP INC  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

12/31/2024 12:49

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.651	0.0	0.070	
ICB1	-1.679	0.0	0.001	
CCV1	236.479	0.0	0.168	
CCB1	-0.868	0.0	0.002	
PB165932BL	-0.340	0.0	0.002	
PB165932BS	94.984	0.0	0.069	
LOWPB165932	9.238	0.0	0.009	
HIGHPB165932	463.104	0.0	0.327	
P5380-01	-0.950	0.0	0.002	
P5380-01DUP	-1.094	0.0	0.002	
P5380-01MS	32.359	0.0	0.025	
P5380-01MSD	32.598	0.0	0.025	
CCV2	241.001	0.0	0.171	
CCB2	-1.290	0.0	0.002	

92% (90-110) 12/31/2024  
92% (90-110) RM

N 14  
Mean 85.728  
SD 137.6015  
CV% 160.51

=====

Calibration results      Aquakem 7.2AQ1      Page:      1

CHEMTECH CONSULTING GROUP INC  
284 Sheffield Street, Mountainside, NJ 07092

12/31/2024 10:50      Reviewed by : RM      Instrument ID : Konelab

-----


Test      Total CN

Accepted      12/31/2024      10:50

Factor      1429  
Bias      0.003

Coeff. of det.      0.999913

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-1.1016	0.0000	-
2	5.0PPBCN	0.005	3.6764	5.0000	-265
3	10PPBCN	0.009	9.4531	10.0000	-55
4	50PPBCN	0.037	49.6443	50.0000	-0.7
5	100PPBCN	0.074	102.7892	100.0000	2.8
6	250PPBCN	0.179	252.0733	250.0000	0.8
7	500PPBCN	0.351	498.4653	500.0000	-0.3

12/31/2024  
RM

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 31 12:30:34 2024

Tue Dec 31 12:43:56 2024

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-1.1016	µg/l	12/31/2024 10:45:11	
5.0PPBCN	A	Total CN	P	3.6764	µg/l	12/31/2024 10:45:12	
10PPBCN	A	Total CN	P	9.4531	µg/l	12/31/2024 10:45:13	
50PPBCN	A	Total CN	P	49.6443	µg/l	12/31/2024 10:45:14	
100PPBCN	A	Total CN	P	102.7892	µg/l	12/31/2024 10:45:15	
250PPBCN	A	Total CN	P	252.0733	µg/l	12/31/2024 10:45:16	
500PPBCN	A	Total CN	P	498.4653	µg/l	12/31/2024 10:45:17	
ICV1	S	Total CN	P	96.6511	µg/l	12/31/2024 12:30:34	
ICB1	S	Total CN	P	-1.6794	µg/l	12/31/2024 12:30:36	
CCV1	S	Total CN	P	236.4786	µg/l	12/31/2024 12:30:38	
CCB1	S	Total CN	P	-0.8679	µg/l	12/31/2024 12:30:40	
PB165932BL	S	Total CN	P	-0.3397	µg/l	12/31/2024 12:30:43	
PB165932BS	S	Total CN	P	94.9845	µg/l	12/31/2024 12:38:07	
LOWPB165932	S	Total CN	P	9.2383	µg/l	12/31/2024 12:38:09	
HIGHPB165932	S	Total CN	P	463.1037	µg/l	12/31/2024 12:38:10	
P5380-01	S	Total CN	P	-0.9499	µg/l	12/31/2024 12:38:12	
P5380-01DUP	S	Total CN	P	-1.0936	µg/l	12/31/2024 12:38:15	
P5380-01MS	S	Total CN	P	32.3592	µg/l	12/31/2024 12:43:50	
P5380-01MSD	S	Total CN	P	32.5977	µg/l	12/31/2024 12:43:51	
CCV2	S	Total CN	P	241.0013	µg/l	12/31/2024 12:43:53	
CCB2	S	Total CN	P	-1.2897	µg/l	12/31/2024 12:43:55	

SOP ID : M9030B-Sulfide-12

SDG No : N/A

Matrix : SOIL

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Block ID : MC-1,

Weigh By : RM

Start Digest Date: 12/30/2024 Time : 08:45 Temp : 70 °C

End Digest Date: 12/30/2024 Time : 10:15 Temp : 70 °C

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Filter paper ID : N/A

Prep Technician Signature: RM

pH Meter ID : N/A

Supervisor Signature: 12

Standardized Name	MLS USED	STD REF. # FROM LOG
LCSS	1.25ML	WP111251
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.5M ZINC ACETATE	5.0ML	WP111004
FORMALDEHYDE	2.0ML	W2725
CONC H2SO4	N/A	M6041
pH Paper 0-14	N/A	W3140
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

12/30/2024 RM

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
P5380-01DUP	TAPIAL3-IDW-SOIL-122024-T1DUP	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
P5380-01MS	TAPIAL3-IDW-SOIL-122024-T1MS	5.03	50	N/A	N/A	N/A	N/A	N/A	N/A
P5380-01MSD	TAPIAL3-IDW-SOIL-122024-T1MSD	5.02	50	N/A	N/A	N/A	N/A	N/A	N/A
PB165915BL	PBS915	5.00	50	N/A	N/A	N/A	N/A	N/A	N/A
PB165915BS	LCS915	5.00	50	N/A	N/A	N/A	N/A	N/A	N/A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : SULFIDE-12-30

WorkList ID : 186892

Department : Distillation

Date : 12-30-2024 08:02:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Solid	Sulfide	Cool 4 deg C	WEST04	N31	12/20/2024	9034

Date/Time 12/30/2024 08:05  
Raw Sample Received by: RIM gws  
Raw Sample Relinquished by: CP (Jm)

Date/Time 12/30/2024  
Raw Sample Received by: CP (Jm)  
Raw Sample Relinquished by: RIM gws



SOP ID : M9012B-Total, Amenable and Reactive Cyanide-20

SDG No : N/A

Start Digest Date: 12/31/2024 Time : 08:50 Temp : 123 °C

Matrix : SOIL

End Digest Date: 12/31/2024 Time : 10:20 Temp : 126 °C

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: 

Weigh By : JP

pH Meter ID : N/A

Supervisor Signature: 12


Standard Name	MLS USED	STD REF. # FROM LOG
LCSS	1.0ML	WP109549
MS/MSD SPIKE SOL.	0.4ML	WP110899
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP108640
50% v/v H2SO4	5.0ML	WP110391
51% w/v MgCL2	2.0ML	WP110390
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
S0	S0	N/A	N/A
S5.0	S5.0	N/A	N/A
S10.0	S10.0	N/A	N/A
S100.0	S100.0	N/A	N/A
S250.0	S250.0	N/A	N/A
S500.0	S500.0	N/A	N/A
ICV	ICV	0.5ML	W3011
ICB	ICB	N/A	N/A
CCV	CCV	N/A	N/A
CCB	CCB	N/A	N/A
Midrange	Midrange	N/A	N/A
HIGHSTD	HIGHSTD	5.0ML	WP110899
LOWSTD	LOWSTD	0.1ML	WP110899

## Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/31/2024 10:35	 / C/C	RM (W/L)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/Nitrite	Comment	Prep Pos
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
P5380-01DUP	TAPIAL3-IDW-SOIL-122024-T1DUP	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
P5380-01MS	TAPIAL3-IDW-SOIL-122024-T1MS	1.02	50	N/A	N/A	N/A	N/A	N/A	N/A
P5380-01MSD	TAPIAL3-IDW-SOIL-122024-T1MSD	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
PB165932BL	PBS932	1.00	50	N/A	N/A	N/A	N/A	N/A	N/A
PB165932BS	LCS932	1.00	50	N/A	N/A	N/A	N/A	N/A	N/A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : cn p5380 s

WorkList ID : 186637

Department : Distillation

Date : 12-26-2024 16:30:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Solid	Cyanide	Cool 4 deg C	WEST04	N31	12/20/2024	9012B

Date/Time12/31/202408:10

Raw Sample Received by:JD WOC

Raw Sample Relinquished by:RM gm

Date/Time12/31/202409:30

Raw Sample Received by:RM gm

Raw Sample Relinquished by:JD WOC

**Instrument ID:** WC PH METER-1

**Daily Analysis Runlog For Sequence/QC Batch ID # LB134101**

Review By	jignesh	Review On	12/27/2024 10:02:08 AM
Supervise By	Iwona	Supervise On	12/27/2024 10:57:33 AM
SubDirectory	LB134101	Test	pH
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3107,W3093,W3094,W3071,W3005,W3072		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	12/27/24 09:37		Jignesh	OK
2	CAL2	CAL2	CAL	12/27/24 09:38		Jignesh	OK
3	CAL3	CAL3	CAL	12/27/24 09:40		Jignesh	OK
4	ICV	ICV	ICV	12/27/24 09:44		Jignesh	OK
5	CCV1	CCV1	CCV	12/27/24 09:45		Jignesh	OK
6	P5380-01	TAPIAL3-IDW-SOIL-1	SAM	12/27/24 09:55		Jignesh	OK
7	P5380-01DUP	TAPIAL3-IDW-SOIL-1	DUP	12/27/24 09:56		Jignesh	OK
8	CCV2	CCV2	CCV	12/27/24 10:00		Jignesh	OK

**Instrument ID:** FLAME

**Daily Analysis Runlog For Sequence/QC Batch ID # LB134107**

Review By	rubina	Review On	12/27/2024 12:04:57 PM
Supervise By	Iwona	Supervise On	12/27/2024 12:07:24 PM
SubDirectory	LB134107	Test	Ignitability
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	P5380-01	TAPIAL3-IDW-SOIL-1	SAM	12/27/24 08:30		rubina	OK
2	P5386-01	MOO-24-00398	SAM	12/27/24 08:38		rubina	OK
3	P5386-02	MOO-24-00398	SAM	12/27/24 08:45		rubina	OK
4	P5386-03	MOO-24-00395-96	SAM	12/27/24 08:52		rubina	OK
5	P5386-04	MOO-24-00395-96	SAM	12/27/24 09:00		rubina	OK
6	P5386-04DUP	MOO-24-00395-96DU	DUP	12/27/24 09:07		rubina	OK

**Instrument ID:** TITRAMETRIC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB134121**

Review By	rubina	Review On	12/30/2024 1:59:54 PM
Supervise By	Iwona	Supervise On	1/2/2025 10:20:12 AM
SubDirectory	LB134121	Test	Sulfide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3105,W3114,W3149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	PB165915BL	PB165915BL	MB	12/30/24 13:20		rubina	OK
2	PB165915BS	PB165915BS	LCS	12/30/24 13:23		rubina	OK
3	P5380-01	TAPIAL3-IDW-SOIL-1	SAM	12/30/24 13:26		rubina	OK
4	P5380-01DUP	TAPIAL3-IDW-SOIL-1	DUP	12/30/24 13:29		rubina	OK
5	P5380-01MS	TAPIAL3-IDW-SOIL-1	MS	12/30/24 13:32		rubina	OK
6	P5380-01MSD	TAPIAL3-IDW-SOIL-1	MSD	12/30/24 13:35		rubina	OK

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB134137**

Review By	rubina	Review On	1/2/2025 8:32:20 AM
Supervise By	Iwona	Supervise On	1/2/2025 10:05:59 AM
SubDirectory	LB134137	Test	Cyanide
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	WP111270,WP111271,WP111272,WP111273,WP111274,WP111275,WP111276		
ICV Standard	W3011		
CCV Standard	WP111271		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP109549		
Chk Standard	WP110103,WP111035,WP111278		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	12/31/24 10:45		rubina	OK
2	5.0PPBCN	5.0PPBCN	CAL2	12/31/24 10:45		rubina	OK
3	10PPBCN	10PPBCN	CAL3	12/31/24 10:45		rubina	OK
4	50PPBCN	50PPBCN	CAL4	12/31/24 10:45		rubina	OK
5	100PPBCN	100PPBCN	CAL5	12/31/24 10:45		rubina	OK
6	250PPBCN	250PPBCN	CAL6	12/31/24 10:45		rubina	OK
7	500PPBCN	500PPBCN	CAL7	12/31/24 10:45		rubina	OK
8	ICV1	ICV1	ICV	12/31/24 12:30		rubina	OK
9	ICB1	ICB1	ICB	12/31/24 12:30		rubina	OK
10	CCV1	CCV1	CCV	12/31/24 12:30		rubina	OK
11	CCB1	CCB1	CCB	12/31/24 12:30		rubina	OK
12	PB165932BL	PB165932BL	MB	12/31/24 12:30		rubina	OK
13	PB165932BS	PB165932BS	LCS	12/31/24 12:38		rubina	OK
14	LOWPB165932	LOWPB165932	SAM	12/31/24 12:38		rubina	OK
15	HIGHPB165932	HIGHPB165932	SAM	12/31/24 12:38		rubina	OK
16	P5380-01	TAPIAL3-IDW-SOIL-1	SAM	12/31/24 12:38		rubina	OK
17	P5380-01DUP	TAPIAL3-IDW-SOIL-1	DUP	12/31/24 12:38		rubina	OK
18	P5380-01MS	TAPIAL3-IDW-SOIL-1	MS	12/31/24 12:43		rubina	OK

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB134137**

Review By	rubina	Review On	1/2/2025 8:32:20 AM
Supervise By	Iwona	Supervise On	1/2/2025 10:05:59 AM
SubDirectory	LB134137	Test	Cyanide

STD. NAME	STD REF.#
ICAL Standard	WP111270,WP111271,WP111272,WP111273,WP111274,WP111275,WP111276
ICV Standard	W3011
CCV Standard	WP111271
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	WP109549
Chk Standard	WP110103,WP111035,WP111278

19	P5380-01MSD	TAPIAL3-IDW-SOIL-1	MSD	12/31/24 12:43		rubina	OK
20	CCV2	CCV2	CCV	12/31/24 12:43		rubina	OK
21	CCB2	CCB2	CCB	12/31/24 12:43		rubina	OK



## Prep Standard - Chemical Standard Summary

**Order ID :** P5380  
**Test :** Cyanide,Ignitability,Percent Solids,pH,Sulfide  
**Prepbatch ID :** PB165915,PB165932,  
**Sequence ID/Qc Batch ID:** LB134101,LB134107,LB134121,LB134137,

**Standard ID :**  
WP108640,WP109549,WP110103,WP110390,WP110391,WP110899,WP111004,WP111035,WP111251,WP111269,WP111270,WP111271,WP111272,WP111273,WP111274,WP111275,WP111276,WP111278,

**Chemical ID :**  
E3657,M5673,M6041,M6121,W1994,W2668,W2725,W2882,W2926,W3001,W3005,W3011,W3019,W3071,W3072,W3093,W3094,W3105,W3107,W3112,W3114,W3138,W3139,W3140,W3149,W3154,

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
11	Sodium hydroxide absorbing solution 0.25 N	<a href="#">WP108640</a>	07/05/2024	01/05/2025	Rubina Mughal	WETCHEM_SCALE_4 (WC SC-4)	None	Iwona Zarych 07/08/2024
<b>FROM</b> 21.00000L of W3112 + 210.00000gram of E3657 = Final Quantity: 21.000 L								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3371	Cyanide LCS Spike Solution, 5PPM	<a href="#">WP109549</a>	09/06/2024	01/05/2025	Niha Farheen Shaik	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 09/06/2024
<b>FROM</b> 1.00000ml of W3138 + 199.00000ml of WP108640 = Final Quantity: 200.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
539	CN BUFFER	<a href="#">WP110103</a>	10/08/2024	04/08/2025	Rubina Mughal	WETCHEM_SCALE_5 (WC SC-5)	None	Iwona Zarych 10/08/2024
<b>FROM</b> 138.00000gram of W2668 + 862.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3214	Magnesium Chloride For Cyanide 2.5M(51%W/V)	<a href="#">WP110390</a>	10/24/2024	04/24/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WC SC-5)	None	Iwona Zarych 10/24/2024
<b>FROM</b> 500.00000ml of W3112 + 510.00000gram of W3001 = Final Quantity: 1000.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1714	Sulfuric Acid, 50% (v/v)	<a href="#">WP110391</a>	10/24/2024	04/24/2025	Niha Farheen Shaik	None	None	Iwona Zarych 10/24/2024

**FROM** 1000.00000ml of M5673 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3850	Cyanide MS-MSD spiking solution, 5PPM	<a href="#">WP110899</a>	12/02/2024	01/05/2025	Iwona Zarych	None	WETCHEM_FIPETTE_3 (WC)	Jignesh Parikh 12/03/2024

**FROM** 1.00000ml of W3154 + 199.00000ml of WP108640 = Final Quantity: 200.000 ml

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
160	0.5M ZINC ACETATE	<a href="#">WP111004</a>	12/09/2024	05/13/2025	Rubina Mughal	WETCHEM_SCALE_8 (WC)	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 12/09/2024
<b>FROM</b> 0.88900L of W3112 + 1.00000ml of M6121 + 110.00000gram of W2926 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
607	PYRIDINE-BARBITURIC ACID	<a href="#">WP111035</a>	12/09/2024	04/30/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WC)	Glass Pipette-A	Iwona Zarych 12/10/2024
<b>FROM</b> 145.00000ml of W3112 + 15.00000gram of W2882 + 15.00000ml of M6121 + 75.00000ml of W3019 = Final Quantity: 250.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3311	Sulfide Int std, 1000PPM	<a href="#">WP111251</a>	12/30/2024	12/31/2024	Rubina Mughal	WETCHEM_SCALE_5 (WC SC-5)	None	Iwona Zarych 01/02/2025
<b>FROM</b> 0.75000gram of W1994 + 99.00000ml of W3112 = Final Quantity: 100.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3456	Cyanide Intermediate Working Std, 5PPM	<a href="#">WP111269</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 0.25000ml of W3154 + 49.75000ml of WP108640 = Final Quantity: 50.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4	Calibration standard 500 ppb	<a href="#">WP111270</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 45.00000ml of WP108640 + 5.00000ml of WP111269 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3761	Calibration-CCV CN Standard 250 ppb	<a href="#">WP111271</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 2.50000ml of WP111269 + 47.50000ml of WP108640 = Final Quantity: 50.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
6	Calibration Standard 100 ppb	<a href="#">WP111272</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 1.00000ml of WP111269 + 49.00000ml of WP108640 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
7	Calibration Standard 50 ppb	<a href="#">WP111273</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 0.50000ml of WP111269 + 49.50000ml of WP108640 = Final Quantity: 50.000 ml								



## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
8	Calibration Standard 10 ppb	<a href="#">WP111274</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 1.00000ml of WP111270 + 49.00000ml of WP108640 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
9	Calibration Standard 5 ppb	<a href="#">WP111275</a>	12/31/2024	01/01/2025	Rubina Mughal	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/02/2025
<b>FROM</b> 0.50000ml of WP111270 + 49.50000ml of WP108640 = Final Quantity: 50.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
167	0 ppb CN calibration std	<a href="#">WP111276</a>	12/31/2024	01/01/2025	Rubina Mughal	None	None	Iwona Zarych
								01/02/2025

**FROM** 50.00000ml of WP108640 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1582	Chloramine T solution, 0.014M	<a href="#">WP111278</a>	12/31/2024	01/01/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	Glass Pipette-A	Iwona Zarych
								01/02/2025

**FROM** 0.08000gram of W3139 + 20.00000ml of W3112 = Final Quantity: 20.000 ml

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	23D2462010	03/20/2028	09/21/2023 / mohan	09/05/2023 / mohan	M5673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	23D2462010	03/20/2028	08/16/2024 / mohan	08/16/2024 / mohan	M6041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	0000275677	05/13/2025	11/13/2024 / Eman	10/13/2024 / Eman	M6121

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3910-1 / Sodium Sulfide, 500 g	WK21A	04/09/2025	04/09/2015 / apatel	04/09/2015 / apatel	W1994

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3818-5 / SODIUM PHOSPHATE, MONOBAS/HYD, CRYST, ACS, 2.5 KG	0000225799	12/03/2025	04/05/2021 / Alexander	02/10/2020 / apatel	W2668

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EMD-FX0410-5 / FORMALDEHYDE SOLUTION 450ML	60045	06/22/2025	08/19/2024 / lwona	06/22/2020 / apatel	W2725

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	EM-BX0035-3 / Barbituric Acid, 100 gms	1.00132.0100	04/30/2025	12/07/2021 / apatel	11/30/2021 / apatel	W2882

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J4296-1 / ZINC ACETATE,DIHYD,CRYST,500G	383058	07/05/2027	07/05/2022 / ketankumar	07/05/2022 / ketankumar	W2926

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	01237-10KG / Magnesium Chloride Hexahydrate ACS 10KG	002251-03319	06/06/2027	01/23/2023 / lwona	06/06/2022 / lwona	W3001

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL13850-1 / Buffer Solution, PH2 (500ml)	4212E45	12/31/2024	01/31/2023 / lwona	01/31/2023 / lwona	W3005

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
EPA	/ ICV-CN	ICV6-400	12/31/2024	01/03/2024 / lwona	02/20/2020 / lwona	W3011

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
SIGMA ALDRICH	270970-1L / Pyridine 1L	SHBQ2113	04/03/2028	04/03/2023 / lwona	04/03/2023 / lwona	W3019

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14455-3 / buffer solution pH 7 yellow	4308H30	07/31/2025	01/02/2024 / JIGNESH	12/06/2023 / lwona	W3071

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14940-1 / Buffer Solution, PH12 (500ml)	2310P21	04/30/2025	01/02/2024 / JIGNESH	12/07/2023 / lwona	W3072

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	566002 / BUFFER PH 7.00 GREEN 1PINT PK6	44001f99	12/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3093

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	1601-1 / PH 10.01 BUFFER,COLOR CD 475ML	4310g83	03/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3094

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL69870-8 / SODIUM THIOSULFATE,0.025N,4LITRE	4403S13	09/30/2025	04/22/2024 / lwona	04/22/2024 / lwona	W3105

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14055-3 / PH 4 BUFFER SOLUTION	AL14055-3	02/27/2026	09/05/2024 / jignesh	05/13/2024 / jignesh	W3107

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL35830-4 / IODINE SOLUTION .025N 1L	2405D89	05/31/2025	07/10/2024 / lwona	07/10/2024 / lwona	W3114

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	LC135457 / Cyanide Standard, 1000 PPM, Second Source	44080060	01/30/2025	09/06/2024 / lwona	08/28/2024 / lwona	W3138

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	JTE494-6 / CHLORAMINE-T BAKER 250GM	10239484	09/09/2029	09/09/2024 / lwona	09/09/2024 / lwona	W3139

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140444 / TEST PAPERS,PH 0-14,.5 SENSI,100PK	10D0142	09/17/2029	09/17/2024 / lwona	09/17/2024 / lwona	W3140

### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL70850-8 / Starch Solution, 4L	4408P62	08/31/2026	10/16/2024 / lwona	10/16/2024 / lwona	W3149

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	RC2543-4 / CYANIDE STD 1000PPM 4OZ	1411J58	05/31/2025	12/02/2024 / lwona	12/02/2024 / lwona	W3154

# Certificate of Analysis



**Date of Release:** 12/6/2013

**Product:** Sodium Sulfide, Nonahydrate GR ACS, Crystals

**Catalog No.:** SX0770 all size codes


**Grade:** Meets ACS Specifications, Meets Reagent Specifications for testing USP/NF monographs

**CAS #:** 1313-84-4

**Country of Origin:** USA

**FW:** 240.18

**Lot No.:** WK21A



Characteristic	Requirement		Results	UOM
	Minimum	Maximum		
Assay (iodometric)	98.0		101.1	%
Ammonium (NH <sub>4</sub> )		0.005	0.003	%
Appearance	Crystals, colorless or only slight yellow color		Crystals, colorless	
Iron	To pass test		Passes	
Sulfite and thiosulfate (as SO <sub>2</sub> )		0.1	0.003	%

Joe Schoellkopf

-----  
Quality Control Manager

This document has been produced electronically and is valid without a signature.



W3071  
Rec 12/6/23

## Certificate of Analysis 12

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4308H30

Product Number: 1551

Manufacture Date: AUG 09, 2023

Expiration Date: JUL 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Yellow liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.002	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (08/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W2918  
W3001  
rec. 06/06/22  
exp. 06/06/27

---

## Chem-Impex International, Inc.

---

**Tel:** (630) 766-2112  
**E-mail:** sales@chemimpex.com  
**Shipping and Correspondence:**  
935 Dillon Drive  
Wood Dale, IL 60191

**Fax:** (630) 766-2218  
**Web site:** www.chemimpex.com  
**Manufacturing site:**  
825 Dillon Drive  
Wood Dale, IL 60191

---

### Certificate of Analysis

<b>Catalogue Number</b>	01237
<b>Product</b>	<b>Magnesium chloride hexahydrate</b>
<b>Lot Number</b>	002251-03319 Magnesium chloride•6H <sub>2</sub> O
<b>CAS Number</b>	7791-18-6
<b>Molecular Formula</b>	MgCl <sub>2</sub> •6H <sub>2</sub> O
<b>Molecular Weight</b>	203.3

---

<b>Appearance</b>	Colorless crystals, very deliquescent
<b>Heavy Metals</b>	< 5 ppm
<b>Anion</b>	Nitrate : < 0.001% Phosphate : < 5 ppm Sulfate : < 0.002%
<b>Cation</b>	Ammonium : < 0.002% Barium : < 0.005% Calcium : 0.0006% Iron : < 5 ppm Manganese : 1.8 ppm Potassium : 0.0006% Sodium : 0.0008% Strontium : 0.0015%
<b>Insoluble material</b>	0.0025%
<b>Assay by titration</b>	100.29%
<b>Grade</b>	ACS reagent
<b>Storage</b>	Store at RT
<b>Country of Origin</b>	India

## ***Certificate of Analysis***

**Catalog Number: 01237**

**Lot Number: 002251-03319**

**Remarks**

See material safety data sheet for additional information

For laboratory use only

**The foregoing is a copy of the Certificate of Analysis as provided by our supplier**



**Bala Kumar**  
**Quality Control Manager**

W3019  
rec 4/3/23

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: [www.sigmaaldrich.com](http://www.sigmaaldrich.com)Email USA: [techserv@sial.com](mailto:techserv@sial.com)Outside USA: [eurtechserv@sial.com](mailto:eurtechserv@sial.com)

## Certificate of Analysis

Product Name:

Pyridine - anhydrous, 99.8%

Product Number:

270970

Batch Number:

SHBQ2113

Brand:

SIAL

CAS Number:

110-86-1

MDL Number:

MFCD00011732

Formula:


C<sub>5</sub>H<sub>5</sub>N

Formula Weight:

79.10 g/mol

Quality Release Date:

15 DEC 2022



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Purity (GC)	≥ 99.75 %	99.99 %
Water (by Karl Fischer)	≤ 0.003 %	0.002 %
Residue on Evaporation	≤ 0.0005 %	< 0.0001 %

  
Larry Coers, Director  
Quality Control  
Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

**RICCA CHEMICAL COMPANY®**

1841 Broad Street  
Pocomoke City, MD 21851  
<http://www.riccachemical.com>  
1-888-GO-RICCA  
[customerservice@riccachemical.com](mailto:customerservice@riccachemical.com)

# Certificate of Analysis

Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C

Lot Number: 2310P21

Product Number: 1615

Manufacture Date: OCT 24, 2023

Expiration Date: APR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	15	20	25	30	35	40
pH	12.35	12.17	11.99	11.78	11.62	11.46

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed


\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	12.005	0.02	186-I-g, 186-II-g, 191d

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1615-1	4 L natural poly	18 months
1615-16	500 mL clear PET-G	18 months
1615-32	1 L natural poly	18 months
1615-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (10/24/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



## Certificate of Analysis

Date of Release: 2/26/2020

Name: Formaldehyde Solution  
GR ACS  
Meets ACS Specifications

Item No: FX0410 all size codes

Lot / Batch No: 60045

Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Assay	36.5	38.0	36.71	%
Chloride (Cl)		5	<5	ppm
Color (APHA)		10	<10	
Form			Passes test	
Heavy metals (as Pb)		5	<5	ppm
Iron (Fe)		5	0.6	ppm
Residue after ignition		0.005	<0.0050	%
Sulfate (SO <sub>4</sub> )		0.002	<0.0020	%
Titrate acid		0.006	<0.0060	meq/g

Heather Sinn,

-----  
Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Millipore Corporation, an affiliate of Merck KGaA, Darmstadt, Germany  
290 Concord Road  
Billerica, MA 01821  
U.S.A

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.





# Certificate of Analysis

## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

**Chemical Formula:** NaOH  
**Molecular Weight:** 40  
**CAS #:** 1310-73-2  
**Appearance:**

**Manufacture Date:** 12/14/2022  
**Expiration Date:** 12/31/2025

**Storage:** Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.





R: 02/20/20  
53

Instructions for QATS Reference Material: *Inorganic ICV Solutions*

For ICP-MS use: dilute the ICV1 concentrate 50-fold with 1% (v/v) nitric acid; pipet 2 mL of the concentrate into a 100 mL volumetric flask and dilute to volume with 1% (v/v) nitric acid.

**ICV5-0415**

For the cold vapor analysis of mercury by AA: dilute the ICV5 concentrate 100-fold with 2% (v/v) nitric acid; pipet 1 mL of the concentrate into a 100 mL volumetric flask and dilute to volume with 2% (v/v) nitric acid. The ICV5 concentrate is prepared in 0.05% (w/v)  $K_2Cr_2O_7$  and 5% (v/v) nitric acid.

**ICV6-0400**

For the analysis of cyanide: dilute the ICV6 concentrate 100-fold with Type II water; pipet 1 mL of the concentrate into a 100 mL volumetric flask and dilute to volume with Type II water. Distill this solution along with the samples before analysis. The cyanide concentrate is prepared from  $K_3Fe(CN)_6$ , Type II water, and 0.1 % sodium hydroxide, and will decompose rapidly if exposed to light.

**NOTE:** USE TYPE II WATER AND HIGH-PURITY ACIDS FOR ALL DILUTIONS.

**(D) CERTIFIED CONCENTRATIONS OF QATS ICV1, ICV5, AND ICV6 SOLUTIONS**

ICV1-1014		
Element	Concentration (µg/L) (after 10-fold dilution)	Concentration (µg/L) (after 50-fold dilution)
Al	2520	504
Sb	1010	202
As	997	199
Ba	518	104
Be	514	103
Cd	514	103
Ca	10000	2000
Cr	517	103
Co	521	104
Cu	505	101
Fe	10100	2020
Pb	1030	206
Mg	5990	1198
Mn	524	105
Ni	525	105
K	9940	1988
Se	1030	206
Ag	252	50
Na	10100	2020
Tl	1040	208
V	504	101
Zn	1010	202

ICV5-0415		ICV6-0400	
Element	Concentration (µg/L) (after 100-fold dilution)	Analyte	Concentration (µg/L) (after 100-fold dilution)
Hg	4.0	CN <sup>-</sup>	99

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

avantor™



Material No.: 9673-33  
Batch No.: 23D2462010  
Manufactured Date: 2023-03-22  
Retest Date: 2028-03-20  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS – Assay (H <sub>2</sub> SO <sub>4</sub> )	95.0 – 98.0 %	96.1 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Substances Reducing Permanganate (as SO <sub>2</sub> )	≤ 2 ppm	< 2 ppm
Ammonium (NH <sub>4</sub> )	≤ 1 ppm	1 ppm
Chloride (Cl)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO <sub>3</sub> )	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities – Aluminum (Al)	≤ 30.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	8.5 ppb
Trace Impurities – Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities – Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities – Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gold (Au)	≤ 10.0 ppb	0.5 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities – Iron (Fe)	≤ 50.0 ppb	1.3 ppb
Trace Impurities – Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities – Magnesium (Mg)	≤ 7.0 ppb	0.8 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities – Nickel (Ni)	≤ 2.0 ppb	0.3 ppb
Trace Impurities – Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se)	≤ 50.0 ppb	< 0.1 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	31.5 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium



Material No.: 9673-33  
Batch No.: 23D2462010

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	5.4 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.4 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

avantor™



M 6041-4b  
MS

Material No.: 9673-33  
Batch No.: 23D2462010  
Manufactured Date: 2023-03-22  
Retest Date: 2028-03-20  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS – Assay (H <sub>2</sub> SO <sub>4</sub> )	95.0 – 98.0 %	96.1 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Substances Reducing Permanganate (as SO <sub>2</sub> )	≤ 2 ppm	< 2 ppm
Ammonium (NH <sub>4</sub> )	≤ 1 ppm	1 ppm
Chloride (Cl)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO <sub>3</sub> )	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities – Aluminum (Al)	≤ 30.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	8.5 ppb
Trace Impurities – Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities – Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities – Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gold (Au)	≤ 10.0 ppb	0.5 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities – Iron (Fe)	≤ 50.0 ppb	1.3 ppb
Trace Impurities – Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities – Magnesium (Mg)	≤ 7.0 ppb	0.8 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities – Nickel (Ni)	≤ 2.0 ppb	0.3 ppb
Trace Impurities – Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se)	≤ 50.0 ppb	< 0.1 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	31.5 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

 **avantor™**




Material No.: 9673-33  
Batch No.: 23D2462010

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	5.4 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.4 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Ethier  
Vice President Global Quality

Hydrochloric Acid, 36.5-38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis



R → 16/13/24  
Met dig

M 6121

Material No.: 9530-33  
Batch No.: 0000275677  
Manufactured Date: 2020/12/16  
Retest Date: 2025/12/15  
Revision No: 1

## Certificate of Analysis

Test	Specification	Result
ACS - Assay (as HCl) (by acid-base titrn)	36.5 - 38.0 %	37.6
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Specific Gravity at 60°/60°F	1.185 - 1.192	1.190
ACS - Bromide (Br)	<= 0.005 %	< 0.005
ACS - Extractable Organic Substances	<= 5 ppm	1
ACS - Free Chlorine (as Cl <sub>2</sub> )	<= 0.5 ppm	< 0.5
Phosphate (PO <sub>4</sub> )	<= 0.05 ppm	< 0.03
Sulfate (SO <sub>4</sub> )	<= 0.5 ppm	< 0.3
Sulfite (SO <sub>3</sub> )	<= 0.8 ppm	0.3
Ammonium (NH <sub>4</sub> )	<= 3 ppm	< 1
Trace Impurities - Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities - Aluminum (Al)	<= 10.0 ppb	< 0.2
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities - Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities - Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	29.7
Trace Impurities - Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC


100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	< 0.2
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	< 1
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.1
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	< 10.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.3
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use  
Product Information (not specifications):  
Appearance (clear, fuming liquid)  
Meets ACS Specifications

Country of Origin: US  
Packaging Site: Phillipsburg Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700





# Certificate of Analysis

1.00132.0000 Barbituric acid for analysis EMSURE®  
Batch N020065932

	Spec. Values		Batch Values	
Assay (acidimetric)	≥ 99	%	99.6	%
Identity (IR-spectrum)	passes test		passes test	
Chloride (Cl)	≤ 40	ppm	≤ 40	ppm
Heavy metals (as Pb)	≤ 50	ppm	≤ 50	ppm
Fe (Iron)	≤ 10	ppm	≤ 10	ppm
Sulfated ash	≤ 0.1	%	≤ 0.1	%
Loss on Drying (105 °C)	≤ 0.1	%	≤ 0.1	%
Suitability as reagent (for cyanide determination)	passes test		passes test	

Date of release (DD.MM.YYYY) 17.04.2020  
Minimum shelf life (DD.MM.YYYY) 30.04.2025

Ioannis Chartomatsidis  
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.

Sodium Phosphate, Monobasic, Monohydrate,  
Crystal  
BAKER ANALYZED® A.C.S. Reagent

(sodium dihydrogen phosphate, monohydrate)



Material No.: 3818-05  
Batch No.: 0000225799  
Manufactured Date: 2018/12/05  
Retest Date: 2025/12/03  
Revision No: 1

## Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay ( $\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$ )	98.0 – 102.0 %	99.5
pH of 5% Solution at 25°C	4.1 – 4.5	4.3
Insoluble Matter	$\leq 0.01$ %	$< 0.01$
Chloride (Cl)	$\leq 5$ ppm	$< 5$
ACS – Sulfate ( $\text{SO}_4$ )	$\leq 0.003$ %	$< 0.003$
Calcium (Ca)	$\leq 0.005$ %	$< 0.005$
Potassium (K)	$\leq 0.01$ %	$< 0.01$
Heavy Metals (as Pb)	$\leq 0.001$ %	$< 0.001$
Trace Impurities – Iron (Fe)	$\leq 0.001$ %	$< 0.001$

For Laboratory, Research or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: IN  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality


For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

## Certificate of Analysis

Product Name:

Zinc acetate dihydrate - ACS reagent, ≥98%

**Product Number:** 383058  
**Batch Number:** MKCQ9159  
**Brand:** SIGALD  
**CAS Number:** 5970-45-6  
**MDL Number:** MFCD00066961  
**Formula:** C<sub>4</sub>H<sub>6</sub>O<sub>4</sub>Zn · 2H<sub>2</sub>O  
**Formula Weight:** 219.51 g/mol  
**Quality Release Date:** 06 JAN 2022



W2926  
 Open 7/5/22  
 received  
 on 7/5/22

Test	Specification	Result
Appearance (Color)	White	White
Appearance (Form)	Powder or Crystal or Chunk(s)	Powder
Infrared Spectrum	Conforms to Structure	Conforms
Insoluble Matter	≤ 0.005 %	0.003 %
Calcium (Ca)	≤ 0.005 %	0.003 %
Chloride (Cl)	≤ 5 ppm	< 5 ppm
Iron (Fe)	≤ 5 ppm	< 5 ppm
Potassium (K)	≤ 0.01 %	0.00 %
Magnesium (Mg)	≤ 0.005 %	0.003 %
Sodium (Na)	≤ 0.05 %	0.03 %
Lead (Pb)	≤ 0.002 %	< 0.001 %
pH	6.0 - 7.0	6.1
Sulfate (SO <sub>4</sub> )	≤ 0.005 %	< 0.005 %
Complexometric EDTA	98.0 - 101.0 %	100.3 %
Meets ACS Requirements	Meets Requirements	Meets Requirements



Larry Coers, Director  
 Quality Control  
 Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





# RICCA CHEMICAL COMPANY®

W 3005

REC- 1/31/23

12

1490 Lammers Pike

Batesville, IN 47006

<http://www.riccachemical.com>

1-888-GO-RICCA

customerservice@riccachemical.com

## Certificate of Analysis

Buffer, Reference Standard, pH 2.00 ± 0.01 at 25°C

Lot Number: 4212E45

Product Number: 1493

Manufacture Date: DEC 20, 2022

Expiration Date: DEC 2024

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	10	15	20	25	30	35	40	45	50
pH	1.93	1.98	1.98	2.00	2.01	2.03	2.03	2.04	2.04

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	2.000	0.02	185i, 186-I-g, 186-II-g

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1493-1	4 L natural poly	24 months
1493-16	500 mL natural poly	24 months
1493-32	1 L natural poly	24 months
1493-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (12/20/2022)

Production Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



## Certificate of Analysis

W3093  
094121  
04/03/2024  
16

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4401F99

Product Number: 1551

Manufacture Date: JAN 08, 2024

Expiration Date: DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	

Test	Specification	Result
Appearance	Yellow liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.004	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-1	4 L natural poly	24 months
1551-1CT	4 L Cubitainer®	24 months
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.





## Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 4310G83

Product Number: 1601

Manufacture Date: OCT 09, 2023

Expiration Date: MAR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	50
pH	10.31	10.23	10.17	10.11	10.05	10.00	9.95	9.91	9.87	9.81

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	Reagent
Preservative	Proprietary	
Blue Dye	Proprietary	

Test	Specification	Result
Appearance	Blue liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	10.003	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
Buffer C	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-16	500 mL natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)





Paul Brandon (10/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

# Certificate of Analysis

**Sodium Thiosulfate, 0.0250 Normal (N/40)**

**Lot Number:** 4403S13

**Product Number:** 7900

**Manufacture Date:** MAR 29, 2024

**Expiration Date:** SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Thiosulfate Pentahydrate	10102-17-7	ACS
Organic Preservative	Proprietary	
Sodium Carbonate	497-19-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.02499-0.02501 N at 20°C	0.02501 N at 20°C	136

Specification	Reference
Standard Sodium Thiosulfate Solution, 0.0250 N	APHA (4500-S2- F)
Standard Sodium Thiosulfate Titrant	APHA (4500-O D)
Standard Sodium Thiosulfate Titrant	APHA (4500-O E)
Standard Sodium Thiosulfate Titrant	APHA (4500-O F)
Standard Sodium Thiosulfate Titrant, 0.025 N	APHA (4500-CI B)
Standard Sodium Thiosulfate Titrant	APHA (4500-O C)
Standard Sodium Thiosulfate Titrant, 0.025 M	APHA (5530 C)
Standard Sodium Thiosulfate Solution (0.025 N)	EPA (SW-846) (9031)
Standard Sodium Thiosulfate solution (0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7900-1	4 L natural poly	18 months
7900-16	500 mL natural poly	18 months
7900-1CT	4 L Cubitainer®	18 months
7900-32	1 L natural poly	18 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



## Certificate of Analysis

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 4403F90

Product Number: 1501

Manufacture Date: MAR 09, 2024

Expiration Date: FEB 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	4.00	4.00	4.00	4.00	4.00	4.00	4.01	4.02	4.03	4.04	4.06

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Preservative	Proprietary	Commercial
Red Dye	Proprietary	Purified

Test	Specification	Result
Appearance	Red liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	4.000	0.02	185i, 186-I-g, 186-II-g

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer B	ASTM (D 5464)
Buffer B	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1501-2.5	10 L Cubitainer®	24 months
1501-32	1 L natural poly	24 months
1501-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Paul Brandon

## Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

# Certificate of Analysis

**Iodine (Iodine-Iodide), 0.0250 Normal (N/40), 1 mL = 0.4008 mg S<sup>2-</sup>**

**Lot Number:** 2405D89

**Product Number:** 3975

**Manufacture Date:** MAY 10, 2024

**Expiration Date:** MAY 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Iodide	7681-11-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	0.02498-0.02502 N at 20°C	0.02502 N at 20°C	136

Specification	Reference
Standard Iodine Solution, 0.0250 N	APHA (4500-S2- F)
Iodine Solution (approximately 0.025 N)	EPA (SW-846) (9031)
Standard Iodine Solution, 0.0250 N	EPA (376.1)
Iodine Solution (approximately 0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3975-1	4 L amber glass	12 months
3975-16	500 mL amber glass	12 months
3975-32	1 L amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (05/10/2024)  
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

## Certificate of Analysis

### Cyanide Standard 1000 ppm (1ml = 1mg CN)

Product Code: **LC13545**

Manufacture Date: August 01, 2024

Lot Number: **44080060**

Expiration Date: January 30, 2025

Test	Specification	Result
Appearance (clarity)	clear solution	clear solution
Appearance (color)	colorless	colorless
Concentration (CN)	0.990 - 1.010mg/mL	1.008mg/mL
Concentration (CN)	990 - 1,010ppm	1,008ppm
Traceable to NIST SRM	Report	999b

**Intended Use** - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

**Storage Information** - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

**Instructions for Handling and Use** - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

**Preparation** - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST standards.

\*The suffix of the product code may differ from what is on your product label. The suffix will designate the size and be associated with a numeric digit(s). Visit [LabChem.com](http://LabChem.com) for more information\*

Suffix	1	2	3/3S/36/36S	4/4C	5	6	7	8	9	20	44	200	246	486
Size	500mL or g	1L or 1kg	2.5L/2.5L Coated/6x2.5L/6x2.5L Coated	4L	20L	10L	125mL	25g	100g	20x20mL	4x4L	200L	24x6mL	48x6mL



Michael Monteleone  
Chemistry Supervisor - Quality Control

ISO9001:2015 Registration #0306-01

W3139 Received on 9/9/24 by IZ

Product No.: A12044  
Product: Chloramine-T trihydrate, 98%  
Lot No.: 10239484

Appearance:	White powder
Melting Point:	166°C(dec)
Assay (Iodometric titration):	100.5%
Identification (FTIR):	Conforms

Order our products online [thermofisher.com/chemicals](https://thermofisher.com/chemicals)

**This document has been electronically generated and does not require a signature.**

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.





# Certificate of Analysis

## Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Starch, soluble	9005-84-9	ACS
Salicylic Acid	69-72-7	ACS

Test	Specification	Result
Appearance	White translucent liquid	Passed
Suitability for Use	Colorless (Iodine absent) - Blue (Iodine present)	Passed

Specification	Reference
Starch Solution	APHA (4500-S2- F)
Starch Indicator Solution	APHA (4500-CI B)
Starch Indicator	APHA (4500-SO32- B)
Starch indicator solution	APHA (2350 B)
Starch indicator solution	APHA (2350 E)
Starch Solution	APHA (510 B)
Starch Solution	APHA (5530 C)
Starch Indicator	APHA (4500-CI C)
Starch Indicator	EPA (345.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8000-1	4 L natural poly	24 months
8000-16	500 mL natural poly	24 months
8000-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

## Production Manager

94 of 102



# Certificate of Analysis

## Cyanide Standard, 1000 ppm CN<sup>-</sup>

**Lot Number:** 1411J58**Product Number:** 2543**Manufacture Date:** NOV 22, 2024**Expiration Date:** MAY 2025

This standard is prepared using accurate volumetric techniques from material that has been assayed against Silver Nitrate solution certified traceable to NIST Standard Reference Material 999. The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is the combined uncertainty based on the stability of the assayed Potassium Cyanide, and the uncertainty in the mass and volume measurements.

Use 0.16% (w/v) (0.04 N) Sodium Hydroxide or 0.225 % (w/v) (0.04 N) Potassium Hydroxide to make dilutions of this standard. Restandardize weekly if extreme accuracy is required.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Cyanide	151-50-8	ACS
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed
Cyanide (CN <sup>-</sup> )	995-1005 ppm	1000 ppm

Specification	Reference
Stock Standard Cyanide Solution	APHA (4500-CN- F)
Stock Cyanide Solution	APHA (4500-CN- E)
Stock Cyanide Solution	APHA (4500-CN- K)
Stock Cyanide Solution	APHA (4500-CN- H)
Cyanide Reference Solution (1000 mg/L)	EPA (SW-846) (7.3.3.2)
Cyanide Calibration Stock Solution (1,000 mg/L CN <sup>-</sup> )	EPA (SW-846) (9213)
Stock Cyanide Solution	EPA (335.3)
Stock Cyanide Solution	EPA (335.2)
Cyanide Solution Stock	ASTM (D 4282)
Simple Cyanide Solution, Stock (1.0 g/L CN <sup>-</sup> )	ASTM (D 4374)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2543-16	500 mL amber poly	6 months
2543-32	1 L amber poly	6 months
2543-4	120 mL amber poly	6 months

**Recommended Storage:** 2°C - 8°C (36°F - 46°F)



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/27/2024

OVENTEMP IN Celsius(°C): 106  
Time IN: 17:15  
In Date: 12/26/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 12/27/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB134080

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	1	1.15	8.48	9.63	8.47	86.3	
P5382-01	COMP-1	2	1.19	8.60	9.79	8.31	82.8	
P5382-02	COMP-2	3	1.15	8.82	9.97	8.43	82.5	
P5382-03	COMP-3	4	1.16	8.76	9.92	8.41	82.8	
P5382-04	SB-1	5	1.15	8.80	9.95	8.47	83.2	
P5382-05	SB-2	6	1.18	8.75	9.93	8.4	82.5	
P5382-06	SB-3	7	1.17	8.80	9.97	8.27	80.7	
P5382-07	SB-4	8	1.19	8.53	9.72	8.02	80.1	
P5382-08	SB-5	9	1.15	8.80	9.95	8.73	86.1	
P5382-09	SB-6	10	1.19	8.50	9.69	7.75	77.2	
P5382-10	SB-7	11	1.14	8.69	9.83	7.86	77.3	
P5382-11	SB-8	12	1.13	8.82	9.95	8.55	84.1	
P5382-12	SB-9	13	1.11	8.73	9.84	8.29	82.2	
P5382-13	SB-10	14	1.19	8.78	9.97	8.16	79.4	
P5382-14	SB-11	15	1.19	8.40	9.59	7.98	80.8	
P5382-15	SB-12	16	1.19	8.51	9.7	8.28	83.3	
P5383-01	OK-02-12232024	17	1.15	8.82	9.97	9.3	92.4	
P5383-02	OK-02-12232024-E2	18	1.13	8.81	9.94	9.54	95.5	
P5384-01	ORA-2066	19	1.00	1.00	2.00	2.00	100.0	wipe sample
P5384-02	ORA-2067	20	1.00	1.00	2.00	2.00	100.0	wipe sample
P5386-01	MOO-24-00398	21	1.15	8.44	9.59	8.99	92.9	
P5386-03	MOO-24-00395-96	22	1.00	1.00	2.00	2.00	100.0	debris
P5387-01	TR-05-122624	23	1.13	8.66	9.79	8.86	89.3	
P5387-02	TR-05-122624-E2	24	1.14	8.80	9.94	8.91	88.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

JB 134020

WorkList Name : %1-122624      WorkList ID : 186590      Department : Wet-Chemistry      Date : 12-26-2024 07:48:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Solid	Percent Solids	Cool 4 deg C	WEST04	N31	12/20/2024	Chemtech -SO
P5382-01	COMP-1	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-02	COMP-2	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-03	COMP-3	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-04	SB-1	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-05	SB-2	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-06	SB-3	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-07	SB-4	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-08	SB-5	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-09	SB-6	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-10	SB-7	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-11	SB-8	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-12	SB-9	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-13	SB-10	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-14	SB-11	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5382-15	SB-12	Solid	Percent Solids	Cool 4 deg C	POWE02	N31	12/20/2024	Chemtech -SO
P5383-01	OK-02-12232024	Solid	Percent Solids	Cool 4 deg C	PSEG05	N31	12/20/2024	Chemtech -SO
P5383-02	OK-02-12232024-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	K31	12/23/2024	Chemtech -SO
P5384-01	ORA-2066	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	12/23/2024	Chemtech -SO
P5384-02	ORA-2067	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	12/26/2024	Chemtech -SO
P5386-01	MOO-24-00398	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	12/26/2024	Chemtech -SO

Date/Time 12/26/24 16:00      Date/Time 12/26/24 17:20

Raw Sample Received by: JB WWC      Raw Sample Received by: JB WWC

Raw Sample Relinquished by: JB WWC      Raw Sample Relinquished by: JB WWC

WORKLIST(Hardcopy Internal Chain)

JB 134080

WorkList Name : %1-122624      WorkList ID : 186590      Department : Wet-Chemistry      Date : 12-26-2024 07:48:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5386-03	MOO-24-00395-96	Solid	Percent Solids	Cool 4 deg C	PSEG03	N31	12/26/2024	Chemtech -SO
P5387-01	TR-05-122624	Solid	Percent Solids	Cool 4 deg C	PSEG05	N41	12/26/2024	Chemtech -SO
P5387-02	TR-05-122624-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	N41	12/26/2024	Chemtech -SO

Date/Time 12/26/24  
Raw Sample Received by: JB WEL  
Raw Sample Relinquished by: JB WEL

Date/Time 12/26/24  
Raw Sample Received by: JB WEL  
Raw Sample Relinquished by: JB WEL



# SHIPPING DOCUMENTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14



P5380



Weston COC ID
Weston_20241220

## Chain of Custody Record/Lab Work Request

Page 1 of 1

<b>Client:</b>	Weston Solutions, Inc.		
<b>Project Manager:</b>	David Sembrot		
<b>Street Address:</b>	1400 Weston Way	<b>City:</b>	West Chester
<b>Phone:</b>	610-314-5456	<b>ST, ZIP:</b>	PA, 19038
<b>e-mail:</b>	david.sembrot@westonsolutions.com		
<b>Sampled By:</b>	Cheyenne Harrington		

<b>Project Name:</b>	Fort Meade RI	<b>Project POC:</b>	Nathan Fretz
<b>PO Number</b>	0111169	<b>Phone:</b>	464-524-5665
<b>W.O. #:</b>		<b>POC e-mail:</b>	nathan.fretz@westonsolutions.com
<b>Lab:</b>	CHEMTECH	<b>Lab POC:</b>	Jordan Hedvat
<b>TAT (days):</b>	7	<b>Lab Phone:</b>	908-728-3144
<b>Lab Address:</b>	284 Sheffield Street Mountainside, NJ 07092		

<b>Matrix Codes</b>
SB- Soil
SE - Sediment
SO - Solid
SL - Sludge
GW - Groundwater
W - Water
O - Oil
A - Air
DS - Drum Solids
DL - Drum Liquids
L - EP/TCLP Leachate
WI - Wipe
X - Other
F - Fish

Lab Use Only		
Temperature of cooler when received (°C)		
COC Tape was present and unbroken on outer package?	Y	N
Samples received in good condition?	Y	N
Labels indicate properly preserved?	Y	N
Received within holding times?	Y	N
Discrepancies between sample labels and COC record?	Y	N

<b>Analyses Requested:</b>	TCLP VOCs by EPA 8260D (1311)	TCLP SVOCs by EPA 8270E (1311)	TCLP Metals by EPA 6010D/7470A	TCLP Pesticides by EPA 8081B	TCLP Herbicides by EPA 8151A	Total Sulfide by EPA 8034	Total Cyanide by EPA 9012E	PCB by EPA 8082A	Ignitability by EPA 1030	pH by EPA 9045D		
<b>Container Type:</b>	Encore	Glass	Glass	Glass	Glass	Glass	Glass	Glass	Glass	Glass		
<b>Container Size:</b>	25g	8 oz	8 oz	8 oz	8 oz	8 oz	8 oz	8 oz	8 oz	8 oz		
<b>Preservative:</b>	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6	Ice to 0-6		

#	Sample ID	G/C	Matrix	# Cont	MS/MSD	Date Collected	Time Collected													Special Instructions/Comments
1	TAPIAL3-IDW-Soil-122024-T1	c	DS	6	no	12/20/2024	14:15	X	X	X	X	X	X	X	X	X	X			expedited 7 day TAT
2																				
3																				
4																				
5																				
6																				
7																				
8																				
9																				
10																				
11																				
12																				

<b>Shipping Airbill Number(s):</b>						<b>Cooler Number:</b>			1	of	1
<b>Relinquished By</b>	<b>Date</b>	<b>Time</b>	<b>Received By</b>	<b>Date</b>	<b>Time</b>	<b>Additional Comments</b>					
1.) <i>S. R. V.</i>	12/20/24	18W	<i>Reen</i>	12/21/24	11:00	QSM 6.0 Compliant					
2.)					2-3	Deliverable Requirements: DoD Level IV report, EnviroData EDD, and ERIS-compatible EDD					
3.)											

**Laboratory Certification**

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488