



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5380

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134101

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	12/27/2024
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	12/27/2024
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	12/27/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5380

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134137

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.097	0.099	98	90-110	12/31/2024
Sample ID:	CCV1						
Cyanide		mg/L	0.24	0.25	96	90-110	12/31/2024
Sample ID:	CCV2						
Cyanide		mg/L	0.24	0.25	96	90-110	12/31/2024