

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134125

Review By	jaswal	Review On	12/30/2024 10:38:51 PM
Supervise By	mohan	Supervise On	12/30/2024 10:39:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP83804,MP83805,MP83806,MP83807,MP83808,MP83809
ICV Standard	MP83810
CCV Standard	MP83812
ICSA Standard	
CRI Standard	MP83814
LCS Standard	
Chk Standard	MP83811,MP83813,MP83815,MP83817

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/30/24 13:56		Mohan	OK
2	S0.2	S0.2	CAL2	12/30/24 13:58		Mohan	OK
3	S2.5	S2.5	CAL3	12/30/24 14:01		Mohan	OK
4	S5	S5	CAL4	12/30/24 14:06		Mohan	OK
5	S7.5	S7.5	CAL5	12/30/24 14:16		Mohan	OK
6	S10	S10	CAL6	12/30/24 14:25		Mohan	OK
7	ICV38	ICV38	ICV	12/30/24 14:28		Mohan	OK
8	ICB38	ICB38	ICB	12/30/24 14:30		Mohan	OK
9	CCV88	CCV88	CCV	12/30/24 14:32		Mohan	OK
10	CCB88	CCB88	CCB	12/30/24 14:35		Mohan	OK
11	CRA	CRA	CRDL	12/30/24 14:37		Mohan	OK
12	HighStd	HighStd	HIGH STD	12/30/24 14:39		Mohan	OK
13	ChkStd	ChkStd	SAM	12/30/24 14:41		Mohan	OK
14	PB165912BL	PB165912BL	MB	12/30/24 14:44		Mohan	OK
15	P5362-02	WC-SOIL-20241219	SAM	12/30/24 14:48		Mohan	OK
16	P5380-02	TAPIAL3-IDW-SOIL-1	SAM	12/30/24 14:51		Mohan	OK
17	P5386-02	MOO-24-00398	SAM	12/30/24 14:53		Mohan	OK
18	P5386-04	MOO-24-00395-96	SAM	12/30/24 15:00		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134125

Review By	jaswal	Review On	12/30/2024 10:38:51 PM
Supervise By	mohan	Supervise On	12/30/2024 10:39:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83804,MP83805,MP83806,MP83807,MP83808,MP83809		
ICV Standard	MP83810		
CCV Standard	MP83812		
ICSA Standard			
CRI Standard	MP83814		
LCS Standard			
Chk Standard	MP83811,MP83813,MP83815,MP83817		

19	P5386-04DUP	MOO-24-00395-96DU	DUP	12/30/24 15:03		Mohan	OK
20	CCV89	CCV89	CCV	12/30/24 15:05		Mohan	OK
21	CCB89	CCB89	CCB	12/30/24 15:07		Mohan	OK
22	P5386-04MS	MOO-24-00395-96MS	MS	12/30/24 15:09		Mohan	OK
23	P5386-04MSD	MOO-24-00395-96MS	MSD	12/30/24 15:12		Mohan	OK
24	PB165858TB	PB165858TB	MB	12/30/24 15:14		Mohan	OK
25	P5386-04L	MOO-24-00395-96L	SD	12/30/24 15:16		Mohan	OK
26	P5386-04A	MOO-24-00395-96A	PS	12/30/24 15:18		Mohan	OK
27	PB165912BS	PB165912BS	LCS	12/30/24 15:22		Mohan	OK
28	CCV90	CCV90	CCV	12/30/24 15:24		Mohan	OK
29	CCB90	CCB90	CCB	12/30/24 15:27		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/30/24 13:54		Kareem	OK
2	S1	S1	CAL2	12/30/24 13:58		Kareem	OK
3	S2	S2	CAL3	12/30/24 14:03		Kareem	OK
4	S3	S3	CAL4	12/30/24 14:07		Kareem	OK
5	S4	S4	CAL5	12/30/24 14:11		Kareem	OK
6	S5	S5	CAL6	12/30/24 14:15		Kareem	OK
7	ICV01	ICV01	ICV	12/30/24 14:20		Kareem	OK
8	LLICV01	LLICV01	LLICV	12/30/24 14:24		Kareem	OK
9	ICB01	ICB01	ICB	12/30/24 14:28		Kareem	OK
10	CRI01	CRI01	CRDL	12/30/24 14:32		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/30/24 14:37		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/30/24 14:45		Kareem	OK
13	CCV01	CCV01	CCV	12/30/24 14:49		Kareem	OK
14	CCB01	CCB01	CCB	12/30/24 14:53		Kareem	OK
15	P5342-08	CHRT26634	SAM	12/30/24 14:58		Kareem	OK
16	P5342-09	HT2651	SAM	12/30/24 15:02		Kareem	OK
17	P5382-01	COMP-1	SAM	12/30/24 15:06		Kareem	OK
18	P5382-02	COMP-2	SAM	12/30/24 15:11		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

19	P5382-03	COMP-3	SAM	12/30/24 15:15		Kareem	OK
20	CCV02	CCV02	CCV	12/30/24 15:19		Kareem	OK
21	CCB02	CCB02	CCB	12/30/24 15:23		Kareem	OK
22	P5362-01	WC-SOIL-20241219	SAM	12/30/24 15:28		Kareem	OK
23	P5362-01DUP	WC-SOIL-20241219D	DUP	12/30/24 15:35		Kareem	OK
24	P5362-01L	WC-SOIL-20241219L	SD	12/30/24 15:39		Kareem	OK
25	P5362-01MS	WC-SOIL-20241219M	MS	12/30/24 15:43	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
26	P5362-01MSD	WC-SOIL-20241219M	MSD	12/30/24 15:47	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
27	P5362-01A	WC-SOIL-20241219A	PS	12/30/24 15:51	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
28	P5386-01	MOO-24-00398	SAM	12/30/24 15:55		Kareem	OK
29	P5386-03	MOO-24-00395-96	SAM	12/30/24 15:59		Kareem	OK
30	P5318-01	AU-06-122024	SAM	12/30/24 16:03		Kareem	OK
31	P5387-01	TR-05-122624	SAM	12/30/24 16:08		Kareem	OK
32	CCV03	CCV03	CCV	12/30/24 16:12		Kareem	OK
33	CCB03	CCB03	CCB	12/30/24 16:16		Kareem	OK
34	PB165863BL	PB165863BL	MB	12/30/24 16:20		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

35	PB165863BS	PB165863BS	LCS	12/30/24 16:25	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
36	P5362-02	WC-SOIL-20241219	SAM	12/30/24 16:29		Kareem	OK
37	P5380-02	TAPIAL3-IDW-SOIL-1	SAM	12/30/24 16:33		Kareem	OK
38	P5386-02	MOO-24-00398	SAM	12/30/24 16:38		Kareem	OK
39	P5386-04	MOO-24-00395-96	SAM	12/30/24 16:42		Kareem	OK
40	P5386-04DUP	MOO-24-00395-96DUP	DUP	12/30/24 16:47		Kareem	OK
41	P5386-04L	MOO-24-00395-96L	SD	12/30/24 16:51		Kareem	OK
42	P5386-04MS	MOO-24-00395-96MS	MS	12/30/24 16:55	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
43	P5386-04MSD	MOO-24-00395-96MSD	MSD	12/30/24 16:59	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
44	CCV04	CCV04	CCV	12/30/24 17:04		Kareem	OK
45	CCB04	CCB04	CCB	12/30/24 17:08		Kareem	OK
46	P5386-04A	MOO-24-00395-96A	PS	12/30/24 17:12	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
47	PB165858TB	PB165858TB	MB	12/30/24 17:16	NOT USE	Kareem	Not Ok
48	PB165887BL	PB165887BL	MB	12/30/24 17:21	NOT USE	Kareem	Not Ok
49	PB165887BS	PB165887BS	LCS	12/30/24 17:25	NOT USE	Kareem	Not Ok
50	P5356-01	STAND-PIPE	SAM	12/30/24 17:29		Kareem	OK
51	P5356-01DUP	STAND-PIPEDUP	DUP	12/30/24 17:33		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

52	P5356-01L	STAND-PIPEL	SD	12/30/24 17:38		Kareem	OK
53	P5356-01MS	STAND-PIPEMS	MS	12/30/24 17:42	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
54	P5356-01MSD	STAND-PIPEMSD	MSD	12/30/24 17:46	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
55	P5356-01A	STAND-PIPEA	PS	12/30/24 17:50	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
56	CCV05	CCV05	CCV	12/30/24 17:54		Kareem	OK
57	CCB05	CCB05	CCB	12/30/24 17:58		Kareem	OK
58	P5377-02DL	GAS-TRE-1114DL	SAM	12/30/24 18:02	Confirm wt/vol	Kareem	OK
59	P5386-05	MOO-24-00397	SAM	12/30/24 18:07		Kareem	OK
60	PB165879BL	PB165879BL	MB	12/30/24 18:11		Kareem	OK
61	PB165879BS	PB165879BS	LCS	12/30/24 18:15	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
62	P5390-01	EO-01-12272024	SAM	12/30/24 18:19		Kareem	OK
63	P5391-01	NB-07-12272024	SAM	12/30/24 18:23		Kareem	OK
64	P5391-01DUP	NB-07-12272024DUP	DUP	12/30/24 18:28		Kareem	OK
65	P5391-01L	NB-07-12272024L	SD	12/30/24 18:32		Kareem	OK
66	P5391-01MS	NB-07-12272024MS	MS	12/30/24 18:36	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

67	P5391-01MSD	NB-07-12272024MSD	MSD	12/30/24 18:40	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
68	CCV06	CCV06	CCV	12/30/24 18:44		Kareem	OK
69	CCB06	CCB06	CCB	12/30/24 18:48		Kareem	OK
70	P5391-01A	NB-07-12272024A	PS	12/30/24 18:53	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
71	PB165917BL	PB165917BL	MB	12/30/24 18:57		Kareem	OK
72	PB165917BS	PB165917BS	LCS	12/30/24 19:01	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
73	LR1	LR1	HIGH STD	12/30/24 19:14		Kareem	OK
74	LR2	LR2	HIGH STD	12/30/24 19:19		Kareem	OK
75	CCV07	CCV07	CCV	12/30/24 19:24		Kareem	OK
76	CCB07	CCB07	CCB	12/30/24 19:31		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134166

Review By	kareem	Review On	1/8/2025 3:32:20 PM
Supervise By	jaswal	Supervise On	1/9/2025 1:56:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83552 MP83553 MP83554 MP83555 MP83556 MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560 MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565 MP8366		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/06/25 12:34		Kareem	OK
2	S1	S1	CAL2	01/06/25 12:39		Kareem	OK
3	S2	S2	CAL3	01/06/25 12:43		Kareem	OK
4	S3	S3	CAL4	01/06/25 12:47		Kareem	OK
5	S4	S4	CAL5	01/06/25 12:51		Kareem	OK
6	S5	S5	CAL6	01/06/25 12:55		Kareem	OK
7	ICV01	ICV01	ICV	01/06/25 13:00		Kareem	OK
8	LLICV01	LLICV01	LLICV	01/06/25 13:13		Kareem	OK
9	ICB01	ICB01	ICB	01/06/25 13:17		Kareem	OK
10	CRI01	CRI01	CRDL	01/06/25 13:21		Kareem	OK
11	ICSA01	ICSA01	ICSA	01/06/25 13:26		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	01/06/25 13:34		Kareem	OK
13	CCV01	CCV01	CCV	01/06/25 13:43		Kareem	OK
14	CCB01	CCB01	CCB	01/06/25 13:53		Kareem	OK
15	P5398-01	341-349	SAM	01/06/25 13:57		Kareem	OK
16	P5398-03	352	SAM	01/06/25 14:01		Kareem	OK
17	P5398-05	352	SAM	01/06/25 14:06		Kareem	OK
18	P5398-02	341-349	SAM	01/06/25 14:11		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134166

Review By	kareem	Review On	1/8/2025 3:32:20 PM
Supervise By	jaswal	Supervise On	1/9/2025 1:56:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83552 MP83553 MP83554 MP83555 MP83556 MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560 MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565 MP8366		

19	CCV02	CCV02	CCV	01/06/25 14:15		Kareem	OK
20	CCB02	CCB02	CCB	01/06/25 14:19		Kareem	OK
21	PB165946BL	PB165946BL	MB	01/06/25 14:23		Kareem	OK
22	PB165946BS	PB165946BS	LCS	01/06/25 14:28	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
23	PB165887BL	PB165887BL	MB	01/06/25 14:32		Kareem	OK
24	PB165887BS	PB165887BS	LCS	01/06/25 14:36	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
25	PB165940BL	PB165940BL	MB	01/06/25 14:40		Kareem	OK
26	PB165940BS	PB165940BS	LCS	01/06/25 14:44	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
27	PB165934TB	PB165934TB	MB	01/06/25 14:48		Kareem	OK
28	PB165858TB	PB165858TB	MB	01/06/25 14:53		Kareem	OK
29	P5398-01DUP	341-349DUP	DUP	01/06/25 14:57		Kareem	OK
30	P5398-01L	341-349L	SD	01/06/25 15:02		Kareem	OK
31	CCV03	CCV03	CCV	01/06/25 15:06		Kareem	OK
32	CCB03	CCB03	CCB	01/06/25 15:10		Kareem	OK
33	P5398-01MS	341-349MS	MS	01/06/25 15:14	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
34	P5398-01MSD	341-349MSD	MSD	01/06/25 15:19	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134166

Review By	kareem	Review On	1/8/2025 3:32:20 PM
Supervise By	jaswal	Supervise On	1/9/2025 1:56:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552 MP83553 MP83554 MP83555 MP83556 MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560 MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565 MP8366

35	P5398-01A	341-349A	PS	01/06/25 15:23	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
36	P5398-02DUP	341-349DUP	DUP	01/06/25 15:27		Kareem	OK
37	P5398-02L	341-349L	SD	01/06/25 15:32		Kareem	OK
38	P5398-02MS	341-349MS	MS	01/06/25 15:36	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
39	P5398-02MSD	341-349MSD	MSD	01/06/25 15:40	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
40	P5398-02A	341-349A	PS	01/06/25 15:44	01ml of m6001 and m6010 were added to 10 ml of sample	Kareem	OK
41	CCV04	CCV04	CCV	01/06/25 15:48		Kareem	OK
42	CCB04	CCB04	CCB	01/06/25 15:58		Kareem	OK
43	LR1	LR1	HIGH STD	01/06/25 16:14		Kareem	OK
44	LR2	LR2	HIGH STD	01/06/25 16:19		Kareem	OK
45	CCV05	CCV05	CCV	01/06/25 16:24		Kareem	OK
46	CCB05	CCB05	CCB	01/06/25 16:28		Kareem	OK