



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions **SDG No.:** P5380
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5380 **SAS No.:** P5380
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	21.1	20.0	105	40 - 160	P	12/30/2024	14:32	LB134129
	Barium	91.3	100	91	40 - 160	P	12/30/2024	14:32	LB134129
	Cadmium	6.07	6.0	101	40 - 160	P	12/30/2024	14:32	LB134129
	Chromium	9.09	10.0	91	40 - 160	P	12/30/2024	14:32	LB134129
	Lead	11.3	12.0	94	40 - 160	P	12/30/2024	14:32	LB134129
	Selenium	17.9	20.0	89	40 - 160	P	12/30/2024	14:32	LB134129
	Silver	10.4	10.0	104	40 - 160	P	12/30/2024	14:32	LB134129
CRA	Mercury	0.19	0.2	97	40 - 160	CV	12/30/2024	14:37	LB134125
CRI01	Arsenic	20.1	20.0	100	40 - 160	P	01/06/2025	13:21	lb134166
	Barium	90.6	100	91	40 - 160	P	01/06/2025	13:21	lb134166
	Cadmium	5.94	6.0	99	40 - 160	P	01/06/2025	13:21	lb134166
	Chromium	9.80	10.0	98	40 - 160	P	01/06/2025	13:21	lb134166
	Lead	12.0	12.0	100	40 - 160	P	01/06/2025	13:21	lb134166
	Selenium	18.6	20.0	93	40 - 160	P	01/06/2025	13:21	lb134166
	Silver	10.2	10.0	102	40 - 160	P	01/06/2025	13:21	lb134166