



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Weston Solutions				SDG No.:	P5380				
Contract:	WEST04	Lab Code:	CHEM	Case No.:	P5380	SAS No.:	P5380			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB38	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/30/2024	14:30	LB134125

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5380</u>								
Contract: <u>WEST04</u>		Lab Code: <u>CHEM</u>		Case No.: <u>P5380</u>			SAS No.: <u>P5380</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB88	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/30/2024	14:35	LB134125
CCB89	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/30/2024	15:07	LB134125
CCB90	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	12/30/2024	15:27	LB134125

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions **SDG No.:** P5380
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5380 **SAS No.:** P5380

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	13:17	lb134166
	Barium	100	+/-100	U	25.0	100	P	01/06/2025	13:17	lb134166
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/06/2025	13:17	lb134166
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	13:17	lb134166
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/06/2025	13:17	lb134166
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	13:17	lb134166
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	13:17	lb134166

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5380</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5380</u>	SAS No.: <u>P5380</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	13:53	lb134166
	Barium	100	+/-100	U	25.0	100	P	01/06/2025	13:53	lb134166
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/06/2025	13:53	lb134166
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	13:53	lb134166
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/06/2025	13:53	lb134166
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	13:53	lb134166
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	13:53	lb134166
CCB02	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	14:19	lb134166
	Barium	100	+/-100	U	25.0	100	P	01/06/2025	14:19	lb134166
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/06/2025	14:19	lb134166
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	14:19	lb134166
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/06/2025	14:19	lb134166
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	14:19	lb134166
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	14:19	lb134166
CCB03	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	15:10	lb134166
	Barium	100	+/-100	U	25.0	100	P	01/06/2025	15:10	lb134166
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/06/2025	15:10	lb134166
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	15:10	lb134166
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/06/2025	15:10	lb134166
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	15:10	lb134166
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	15:10	lb134166
CCB04	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	15:58	lb134166
	Barium	100	+/-100	U	25.0	100	P	01/06/2025	15:58	lb134166
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/06/2025	15:58	lb134166
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	15:58	lb134166
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/06/2025	15:58	lb134166
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	15:58	lb134166
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	15:58	lb134166
CCB05	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	16:28	lb134166
	Barium	100	+/-100	U	25.0	100	P	01/06/2025	16:28	lb134166
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/06/2025	16:28	lb134166
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	16:28	lb134166
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/06/2025	16:28	lb134166
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/06/2025	16:28	lb134166
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/06/2025	16:28	lb134166

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5380</u>								
Contract: <u>WEST04</u>		Lab Code: <u>CHEM</u>		Case No.: <u>P5380</u>			SAS No.: <u>P5380</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	14:28	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	14:28	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	14:28	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	14:28	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	14:28	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	14:28	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	14:28	LB134129

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5380</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5380</u>	SAS No.: <u>P5380</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	14:53	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	14:53	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	14:53	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	14:53	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	14:53	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	14:53	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	14:53	LB134129
CCB02	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	15:23	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	15:23	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	15:23	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	15:23	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	15:23	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	15:23	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	15:23	LB134129
CCB03	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	16:16	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	16:16	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	16:16	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	16:16	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	16:16	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	16:16	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	16:16	LB134129
CCB04	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	17:08	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	17:08	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	17:08	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	17:08	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	17:08	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	17:08	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	17:08	LB134129
CCB05	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	17:58	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	17:58	LB134129
	Cadmium	0.24	+/-6.00	J	1.50	6.00	P	12/30/2024	17:58	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	17:58	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	17:58	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	17:58	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	17:58	LB134129
CCB06	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	18:48	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	18:48	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	18:48	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	18:48	LB134129

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions **SDG No.:** P5380
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5380 **SAS No.:** P5380

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	18:48	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	18:48	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	18:48	LB134129
CCB07	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	19:31	LB134129
	Barium	100	+/-100	U	25.0	100	P	12/30/2024	19:31	LB134129
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	12/30/2024	19:31	LB134129
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	19:31	LB134129
	Lead	12.0	+/-12.0	U	9.60	12.0	P	12/30/2024	19:31	LB134129
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	12/30/2024	19:31	LB134129
	Silver	10.0	+/-10.0	U	5.00	10.0	P	12/30/2024	19:31	LB134129