

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

Continuing Calib Date: 12/26/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 19:56 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.80	4.81	4.71	4.91	0.01
Aroclor-1016-2	(2)	4.82	4.83	4.73	4.93	0.01
Aroclor-1016-3	(3)	4.87	4.89	4.79	4.99	0.02
Aroclor-1016-4	(4)	5.00	5.01	4.91	5.11	0.02
Aroclor-1016-5	(5)	5.25	5.27	5.17	5.37	0.02
Aroclor-1260-1	(1)	6.29	6.31	6.21	6.41	0.02
Aroclor-1260-2	(2)	6.48	6.50	6.40	6.60	0.02
Aroclor-1260-3	(3)	6.85	6.87	6.77	6.97	0.02
Aroclor-1260-4	(4)	7.11	7.13	7.03	7.23	0.02
Aroclor-1260-5	(5)	7.35	7.37	7.27	7.47	0.02
Tetrachloro-m-xylene		3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl		8.77	8.79	8.69	8.89	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

Continuing Calib Date: 12/26/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 19:56 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.79	4.80	4.70	4.90	0.02
Aroclor-1016-2	(2)	4.81	4.82	4.72	4.92	0.02
Aroclor-1016-3	(3)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-4	(4)	5.02	5.04	4.94	5.14	0.02
Aroclor-1016-5	(5)	5.24	5.25	5.15	5.35	0.01
Aroclor-1260-1	(1)	6.27	6.29	6.19	6.39	0.02
Aroclor-1260-2	(2)	6.46	6.47	6.37	6.57	0.01
Aroclor-1260-3	(3)	6.61	6.63	6.53	6.73	0.02
Aroclor-1260-4	(4)	7.08	7.10	7.00	7.20	0.02
Aroclor-1260-5	(5)	7.32	7.34	7.24	7.44	0.02
Tetrachloro-m-xylene		3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl		8.72	8.74	8.64	8.84	0.02