

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/26/2024

Lab Sample No.: AR1660CCC500 Data File : PO108758.D Time Analyzed: 09:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.798	4.709	4.909	492.260	500.000	-1.5
Aroclor-1016-2	4.819	4.729	4.929	555.520	500.000	11.1
Aroclor-1016-3	4.875	4.785	4.985	566.500	500.000	13.3
Aroclor-1016-4	4.996	4.907	5.107	554.660	500.000	10.9
Aroclor-1016-5	5.254	5.165	5.365	545.160	500.000	9.0
Aroclor-1260-1	6.296	6.210	6.410	532.910	500.000	6.6
Aroclor-1260-2	6.484	6.398	6.598	520.070	500.000	4.0
Aroclor-1260-3	6.854	6.769	6.969	508.280	500.000	1.7
Aroclor-1260-4	7.114	7.029	7.229	500.340	500.000	0.1
Aroclor-1260-5	7.355	7.270	7.470	505.130	500.000	1.0
Decachlorobiphenyl	8.768	8.691	8.891	48.400	50.000	-3.2
Tetrachloro-m-xylene	3.703	3.610	3.810	58.440	50.000	16.9

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/26/2024

Lab Sample No.: AR1660CCC500 Data File : PO108758.D Time Analyzed: 09:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.785	4.698	4.898	523.530	500.000	4.7
Aroclor-1016-2	4.805	4.718	4.918	530.700	500.000	6.1
Aroclor-1016-3	4.981	4.894	5.094	521.170	500.000	4.2
Aroclor-1016-4	5.022	4.935	5.135	517.040	500.000	3.4
Aroclor-1016-5	5.236	5.150	5.350	538.660	500.000	7.7
Aroclor-1260-1	6.269	6.186	6.386	539.930	500.000	8.0
Aroclor-1260-2	6.456	6.373	6.573	533.550	500.000	6.7
Aroclor-1260-3	6.610	6.527	6.727	525.430	500.000	5.1
Aroclor-1260-4	7.081	7.000	7.200	541.640	500.000	8.3
Aroclor-1260-5	7.321	7.239	7.439	544.970	500.000	9.0
Decachlorobiphenyl	8.717	8.641	8.841	51.720	50.000	3.4
Tetrachloro-m-xylene	3.700	3.608	3.808	54.500	50.000	9.0