

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/26/2024

Lab Sample No.: AR1660CCC500 Data File : PO108773.D Time Analyzed: 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.798	4.709	4.909	500.920	500.000	0.2
Aroclor-1016-2	4.819	4.729	4.929	536.280	500.000	7.3
Aroclor-1016-3	4.874	4.785	4.985	543.260	500.000	8.7
Aroclor-1016-4	4.996	4.907	5.107	553.060	500.000	10.6
Aroclor-1016-5	5.254	5.165	5.365	531.340	500.000	6.3
Aroclor-1260-1	6.296	6.210	6.410	460.940	500.000	-7.8
Aroclor-1260-2	6.484	6.398	6.598	500.470	500.000	0.1
Aroclor-1260-3	6.853	6.769	6.969	489.770	500.000	-2.0
Aroclor-1260-4	7.113	7.029	7.229	448.450	500.000	-10.3
Aroclor-1260-5	7.354	7.270	7.470	434.300	500.000	-13.1
Decachlorobiphenyl	8.766	8.691	8.891	41.820	50.000	-16.4
Tetrachloro-m-xylene	3.703	3.610	3.810	55.860	50.000	11.7

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/26/2024

Lab Sample No.: AR1660CCC500 Data File : PO108773.D Time Analyzed: 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.785	4.698	4.898	546.650	500.000	9.3
Aroclor-1016-2	4.806	4.718	4.918	538.450	500.000	7.7
Aroclor-1016-3	4.981	4.894	5.094	528.610	500.000	5.7
Aroclor-1016-4	5.022	4.935	5.135	515.360	500.000	3.1
Aroclor-1016-5	5.236	5.150	5.350	549.390	500.000	9.9
Aroclor-1260-1	6.270	6.186	6.386	508.560	500.000	1.7
Aroclor-1260-2	6.457	6.373	6.573	501.550	500.000	0.3
Aroclor-1260-3	6.610	6.527	6.727	500.950	500.000	0.2
Aroclor-1260-4	7.083	7.000	7.200	502.140	500.000	0.4
Aroclor-1260-5	7.322	7.239	7.439	505.850	500.000	1.2
Decachlorobiphenyl	8.716	8.641	8.841	47.280	50.000	-5.4
Tetrachloro-m-xylene	3.700	3.608	3.808	53.750	50.000	7.5