

Surrogate Summary

SDG No.: **P5380**

Client: **Weston Solutions**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108361.D	PIBLK-PO108361.D	Tetrachloro-m-xylene	1	20	23.0	115		60	140
		Decachlorobiphenyl	1	20	23.6	118		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PO108762.D	PIBLK-PO108762.D	Tetrachloro-m-xylene	1	20	21.8	109		60	140
		Decachlorobiphenyl	1	20	21.3	107		60	140
		Tetrachloro-m-xylene	2	20	20.0	100		60	140
		Decachlorobiphenyl	2	20	22.1	110		60	140
PB165843BL	PB165843BL	Tetrachloro-m-xylene	1	20	22.0	110		44	130
		Decachlorobiphenyl	1	20	22.2	111		60	125
		Tetrachloro-m-xylene	2	20	20.0	100		44	130
		Decachlorobiphenyl	2	20	22.9	115		60	125
PB165843BS	PB165843BS	Tetrachloro-m-xylene	1	20	21.2	106		44	130
		Decachlorobiphenyl	1	20	21.7	108		60	125
		Tetrachloro-m-xylene	2	20	19.2	96		44	130
		Decachlorobiphenyl	2	20	23.2	116		60	125
P5318-01MS	AU-06-122024MS	Tetrachloro-m-xylene	1	20	13.5	67		44	130
		Decachlorobiphenyl	1	20	10.9	55	*	60	125
		Tetrachloro-m-xylene	2	20	12.4	62		44	130
		Decachlorobiphenyl	2	20	12.1	61		60	125
P5318-01MSD	AU-06-122024MSD	Tetrachloro-m-xylene	1	20	13.9	70		44	130
		Decachlorobiphenyl	1	20	11.8	59	*	60	125
		Tetrachloro-m-xylene	2	20	13.6	68		44	130
		Decachlorobiphenyl	2	20	13.3	66		60	125
P5380-01	TAPIAL3-IDW-SOIL-122024-T1	Tetrachloro-m-xylene	1	20	20.5	102		44	130
		Decachlorobiphenyl	1	20	14.3	71		60	125
		Tetrachloro-m-xylene	2	20	19.4	97		44	130
		Decachlorobiphenyl	2	20	16.1	80		60	125
I.BLK-PO108777.D	PIBLK-PO108777.D	Tetrachloro-m-xylene	1	20	20.5	102		60	140
		Decachlorobiphenyl	1	20	18.1	90		60	140
		Tetrachloro-m-xylene	2	20	19.4	97		60	140
		Decachlorobiphenyl	2	20	19.9	100		60	140