

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5380

Client: Weston Solutions

Analytical Method: 8082A

DataFile : PO108767.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: P5318-01MS	AU-06-122024MS											
	AR1016	193.2	0	138	ug/kg	71				47	134	
	AR1260	193.2	8.7	124	ug/kg	60				53	140	