

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

Instrument ID: ECD\_S Calibration Date(s): 12/23/2024 12/23/2024  
Calibration Times: 11:23 12:59

GC Column: RTX-CLP ID: 0.32 (mm)

| LAB FILE ID:     |             | CF 200 =          | <u>PS028787.D</u> | CF 500 =          | <u>PS028788.D</u> |                   |          |
|------------------|-------------|-------------------|-------------------|-------------------|-------------------|-------------------|----------|
| CF 750 =         |             | <u>PS028789.D</u> | CF 1000 =         | <u>PS028790.D</u> | CF 1500 =         | <u>PS028791.D</u> |          |
| COMPOUND         | CF 200      | CF 500            | CF 750            | CF 1000           | CF 1500           | CF                | %<br>RSD |
| 2,4,5-TP(Silvex) | 16812700000 | 15996100000       | 15801400000       | 15298000000       | 14638400000       | 15709300000       | 5        |
| 2,4-D            | 3146520000  | 2896320000        | 2842430000        | 2725340000        | 2637060000        | 2849530000        | 7        |
| 2,4-DCAA         | 2519360000  | 2280310000        | 2254700000        | 2149750000        | 2106010000        | 2262030000        | 7        |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

Instrument ID: ECD\_S Calibration Date(s): 12/23/2024 12/23/2024  
Calibration Times: 11:23 12:59

GC Column: RTX-CLP2 ID: 0.32 (mm)

| LAB FILE ID:     |            | CF 200 =          | <u>PS028787.D</u> | CF 500 =          | <u>PS028788.D</u> |                   |          |
|------------------|------------|-------------------|-------------------|-------------------|-------------------|-------------------|----------|
| CF 750 =         |            | <u>PS028789.D</u> | CF 1000 =         | <u>PS028790.D</u> | CF 1500 =         | <u>PS028791.D</u> |          |
| COMPOUND         | CF 200     | CF 500            | CF 750            | CF 1000           | CF 1500           | CF                | %<br>RSD |
| 2,4,5-TP(Silvex) | 9147430000 | 9053780000        | 9109830000        | 8920050000        | 8686470000        | 8983510000        | 2        |
| 2,4-D            | 1472990000 | 1449810000        | 1452790000        | 1429890000        | 1410250000        | 1443150000        | 2        |
| 2,4-DCAA         | 1184820000 | 1126220000        | 1123240000        | 1104090000        | 1089310000        | 1125540000        | 3        |