

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** P5380 **SAS No.:** P5380 **SDG NO.:** P5380

Continuing Calib Date: 12/27/2024 **Initial Calibration Date(s):** 12/23/2024 12/23/2024

Continuing Calib Time: 19:56 **Initial Calibration Time(s):** 11:23 12:59

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.21	7.21	7.11	7.31	0.01
2,4-D	8.33	8.33	8.23	8.43	0.01
2,4,5-TP(Silvex)	9.20	9.21	9.11	9.31	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** P5380 **SAS No.:** P5380 **SDG NO.:** P5380

Continuing Calib Date: 12/27/2024 **Initial Calibration Date(s):** 12/23/2024 12/23/2024

Continuing Calib Time: 19:56 **Initial Calibration Time(s):** 11:23 12:59

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.69	7.70	7.60	7.80	0.01
2,4-D	8.93	8.94	8.84	9.04	0.01
2,4,5-TP(Silvex)	9.83	9.84	9.74	9.94	0.01