

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** WEST04

**Lab Code:** CHEM      **Case No.:** P5380      **SAS No.:** P5380      **SDG NO.:** P5380

**Continuing Calib Date:** 12/30/2024      **Initial Calibration Date(s):** 12/23/2024      12/23/2024

**Continuing Calib Time:** 12:02      **Initial Calibration Time(s):** 11:23      12:59

**GC Column:** RTX-CLP      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.21       | 7.21      | 7.11      | 7.31 | 0.00       |
| 2,4-D            | 8.33       | 8.33      | 8.23      | 8.43 | 0.00       |
| 2,4,5-TP(Silvex) | 9.20       | 9.21      | 9.11      | 9.31 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** WEST04

**Lab Code:** CHEM      **Case No.:** P5380      **SAS No.:** P5380      **SDG NO.:** P5380

**Continuing Calib Date:** 12/30/2024      **Initial Calibration Date(s):** 12/23/2024      12/23/2024

**Continuing Calib Time:** 12:02      **Initial Calibration Time(s):** 11:23      12:59

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.69       | 7.70      | 7.60      | 7.80 | 0.01       |
| 2,4-D            | 8.93       | 8.94      | 8.84      | 9.04 | 0.01       |
| 2,4,5-TP(Silvex) | 9.83       | 9.84      | 9.74      | 9.94 | 0.01       |