

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/27/2024

Lab Sample No.: HSTDCCC750 Data File : PS028837.D Time Analyzed: 19:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.198	9.106	9.306	789.440	712.500	10.8
2,4-D	8.325	8.232	8.432	737.600	705.000	4.6
2,4-DCAA	7.205	7.110	7.310	805.920	750.000	7.5

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/27/2024

Lab Sample No.: HSTDCCC750 Data File : PS028837.D Time Analyzed: 19:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.834	9.742	9.942	729.080	712.500	2.3
2,4-D	8.932	8.840	9.040	705.540	705.000	0.1
2,4-DCAA	7.694	7.600	7.800	754.200	750.000	0.6