

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** P5380 **SAS No.:** P5380 **SDG NO.:** P5380

GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 12/23/2024 12/23/2024

Client Sample No.: CCAL03 **Date Analyzed:** 12/30/2024

Lab Sample No.: HSTDCCC750 **Data File :** PS028849.D **Time Analyzed:** 10:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.204	9.106	9.306	757.550	712.500	6.3
2,4-D	8.330	8.232	8.432	734.230	705.000	4.1
2,4-DCAA	7.210	7.110	7.310	785.200	750.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL03 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028849.D Time Analyzed: 10:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.833	9.742	9.942	659.120	712.500	-7.5
2,4-D	8.932	8.840	9.040	632.820	705.000	-10.2
2,4-DCAA	7.694	7.600	7.800	675.500	750.000	-9.9