

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL04 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028853.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.203	9.106	9.306	753.330	712.500	5.7
2,4-D	8.329	8.232	8.432	732.860	705.000	4.0
2,4-DCAA	7.209	7.110	7.310	778.690	750.000	3.8

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL04 Date Analyzed: 12/30/2024

Lab Sample No.: HSTDCCC750 Data File : PS028853.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.833	9.742	9.942	662.660	712.500	-7.0
2,4-D	8.932	8.840	9.040	630.920	705.000	-10.5
2,4-DCAA	7.694	7.600	7.800	679.980	750.000	-9.3