

### CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

Continuing Calib Date: 12/27/2024 Initial Calibration Date(s): 12/23/2024 12/23/2024

Continuing Calib Time: 14:32 Initial Calibration Time(s): 13:15 14:09

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.05       | 9.06      | 8.96                   | 9.16 | 0.01       |
| Tetrachloro-m-xylene | 3.54       | 3.54      | 3.44                   | 3.64 | 0.00       |
| gamma-BHC (Lindane)  | 4.33       | 4.33      | 4.23                   | 4.43 | 0.00       |
| Heptachlor           | 4.92       | 4.92      | 4.82                   | 5.02 | 0.00       |
| Heptachlor epoxide   | 5.68       | 5.69      | 5.59                   | 5.79 | 0.01       |
| Endrin               | 6.57       | 6.58      | 6.48                   | 6.68 | 0.01       |
| Methoxychlor         | 7.50       | 7.50      | 7.40                   | 7.60 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** WEST04

**Lab Code:** CHEM      **Case No.:** P5380      **SAS No.:** P5380      **SDG NO.:** P5380

**Continuing Calib Date:** 12/27/2024      **Initial Calibration Date(s):** 12/23/2024      12/23/2024

**Continuing Calib Time:** 14:32      **Initial Calibration Time(s):** 13:15      14:09

**GC Column:** ZB-MR2      **ID:** 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Decachlorobiphenyl   | 7.91       | 7.91      | 7.81      | 8.01 | 0.00       |
| Tetrachloro-m-xylene | 2.77       | 2.78      | 2.68      | 2.88 | 0.01       |
| gamma-BHC (Lindane)  | 3.61       | 3.61      | 3.51      | 3.71 | 0.00       |
| Heptachlor           | 3.95       | 3.95      | 3.85      | 4.05 | 0.00       |
| Heptachlor epoxide   | 4.73       | 4.73      | 4.63      | 4.83 | 0.00       |
| Endrin               | 5.64       | 5.64      | 5.54      | 5.74 | 0.00       |
| Methoxychlor         | 6.61       | 6.61      | 6.51      | 6.71 | 0.00       |