

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/27/2024

Lab Sample No.: PSTDCCC050 Data File : PL093554.D Time Analyzed: 18:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.054	8.956	9.156	50.540	50.000	1.1
Endrin	6.573	6.475	6.675	44.470	50.000	-11.1
gamma-BHC (Lindane)	4.328	4.229	4.429	51.230	50.000	2.5
Heptachlor	4.916	4.818	5.018	49.360	50.000	-1.3
Heptachlor epoxide	5.684	5.586	5.786	50.190	50.000	0.4
Methoxychlor	7.500	7.400	7.600	47.780	50.000	-4.4
Tetrachloro-m-xylene	3.539	3.442	3.642	50.770	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG NO.: P5380

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/27/2024

Lab Sample No.: PSTDCCC050 Data File : PL093554.D Time Analyzed: 18:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.912	7.812	8.012	53.360	50.000	6.7
Endrin	5.639	5.540	5.740	47.950	50.000	-4.1
gamma-BHC (Lindane)	3.609	3.510	3.710	52.940	50.000	5.9
Heptachlor	3.947	3.848	4.048	50.970	50.000	1.9
Heptachlor epoxide	4.729	4.630	4.830	51.550	50.000	3.1
Methoxychlor	6.612	6.512	6.712	48.040	50.000	-3.9
Tetrachloro-m-xylene	2.776	2.677	2.877	52.580	50.000	5.2