

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122724\
 Data File : PL093541.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2024 14:32
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 00:57:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122324.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 24 15:29:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.539	2.774	130.2E6	157.5E6	52.589	54.117
2)	SA Decachlor...	9.054	7.912	95503809	158.6E6	51.651	53.115
Target Compounds							
2)	A alpha-BHC	3.995	3.277	183.4E6	240.9E6	53.127	55.422
3)	MA gamma-BHC...	4.327	3.608	174.7E6	233.8E6	53.260	55.407
4)	MA Heptachlor	4.916	3.946	153.2E6	227.3E6	52.331	54.692
5)	MB Aldrin	5.257	4.226	151.9E6	226.7E6	52.224	55.274
6)	B beta-BHC	4.525	3.908	76370661	98948563	52.977	55.048
7)	B delta-BHC	4.773	4.136	166.7E6	239.3E6	54.435	56.577
8)	B Heptachlo...	5.684	4.729	137.9E6	207.4E6	52.338	54.169
9)	A Endosulfan I	6.070	5.098	122.1E6	180.9E6	51.751	51.765
10)	B gamma-Chl...	5.940	4.979	130.3E6	211.6E6	51.855	54.907
11)	B alpha-Chl...	6.019	5.042	130.6E6	207.3E6	52.167	54.438
12)	B 4,4'-DDE	6.193	5.231	119.4E6	205.6E6	53.200	55.915
13)	MA Dieldrin	6.345	5.362	129.2E6	210.2E6	51.791	54.547
14)	MA Endrin	6.574	5.638	112.6E6	181.6E6	52.315	54.878
15)	B Endosulfa...	6.794	5.933	112.4E6	180.2E6	49.462	55.479
16)	A 4,4'-DDD	6.709	5.786	97729958	163.3E6	55.656	57.716
17)	MA 4,4'-DDT	7.024	6.037	96544968	164.3E6	52.228	54.401
18)	B Endrin al...	6.924	6.112	91177159	145.0E6	51.381	53.831
19)	B Endosulfa...	7.158	6.335	104.6E6	172.0E6	51.815	54.521
20)	A Methoxychlor	7.500	6.611	53535568	87904365	53.553	54.611
21)	B Endrin ke...	7.643	6.840	116.9E6	196.8E6	52.110	54.048
22)	Mirex	8.117	7.020	94046492	156.8E6	50.327	51.301
24)	Chlordane-2	5.257f	4.365	151.9E6	5474380	1381.963	39.438 #
25)	Chlordane-3	5.940	4.979	130.3E6	211.6E6	342.869	498.535 #
26)	Chlordane-4	6.019	5.042	130.6E6	207.3E6	287.382	503.881 #
27)	Chlordane-5	0.000	5.933	0	180.2E6	N.D.	1353.790 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122724\
Data File : PL093541.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2024 14:32
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 00:57:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122324.M
Quant Title : GC Extractables
QLast Update : Tue Dec 24 15:29:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.539 min
Delta R.T.: -0.003 min
Response: 130194258
Conc: 52.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: -0.003 min
Response: 157545989
Conc: 54.12 ng/ml




#2 alpha-BHC

R.T.: 3.995 min
Delta R.T.: -0.002 min
Response: 183415347
Conc: 53.13 ng/ml



#2 alpha-BHC


R.T.: 3.277 min
Delta R.T.: -0.003 min
Response: 240906509
Conc: 55.42 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.327 min
Delta R.T.: -0.002 min
Response: 174661804
Conc: 53.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.608 min
Delta R.T.: -0.002 min
Response: 233783062
Conc: 55.41 ng/ml




#4 Heptachlor

R.T.: 4.916 min
Delta R.T.: -0.002 min
Response: 153241993
Conc: 52.33 ng/ml



#4 Heptachlor


R.T.: 3.946 min
Delta R.T.: -0.002 min
Response: 227296779
Conc: 54.69 ng/ml



#5 Aldrin


R.T.: 5.257 min
Delta R.T.: -0.002 min
Response: 151911125
Conc: 52.22 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.226 min
Delta R.T.: -0.002 min
Response: 226749023
Conc: 55.27 ng/ml




#6 beta-BHC

R.T.: 4.525 min
Delta R.T.: -0.002 min
Response: 76370661
Conc: 52.98 ng/ml



#6 beta-BHC


R.T.: 3.908 min
Delta R.T.: -0.002 min
Response: 98948563
Conc: 55.05 ng/ml



#7 delta-BHC


R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 166712800
Conc: 54.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.136 min
Delta R.T.: -0.002 min
Response: 239255807
Conc: 56.58 ng/ml




#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: -0.002 min
Response: 137877438
Conc: 52.34 ng/ml



#8 Heptachlor epoxide


R.T.: 4.729 min
Delta R.T.: -0.002 min
Response: 207396466
Conc: 54.17 ng/ml



#9 Endosulfan I


R.T.: 6.070 min
Delta R.T.: -0.001 min
Response: 122092883
Conc: 51.75 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.098 min
Delta R.T.: -0.002 min
Response: 180864778
Conc: 51.77 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: -0.001 min
Response: 130315187
Conc: 51.85 ng/ml



#10 gamma-Chlordane


R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 211556858
Conc: 54.91 ng/ml



#11 alpha-Chlordane


R.T.: 6.019 min
Delta R.T.: -0.002 min
Response: 130570041
Conc: 52.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.042 min
Delta R.T.: -0.002 min
Response: 207265394
Conc: 54.44 ng/ml




#12 4,4'-DDE

R.T.: 6.193 min
Delta R.T.: -0.001 min
Response: 119373643
Conc: 53.20 ng/ml



#12 4,4'-DDE


R.T.: 5.231 min
Delta R.T.: -0.002 min
Response: 205613552
Conc: 55.92 ng/ml



#13 Dieldrin


R.T.: 6.345 min
Delta R.T.: -0.001 min
Response: 129222423
Conc: 51.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.362 min
Delta R.T.: -0.002 min
Response: 210217206
Conc: 54.55 ng/ml




#14 Endrin

R.T.: 6.574 min
Delta R.T.: -0.001 min
Response: 112605727
Conc: 52.32 ng/ml



#14 Endrin


R.T.: 5.638 min
Delta R.T.: -0.002 min
Response: 181554809
Conc: 54.88 ng/ml



#15 Endosulfan II


R.T.: 6.794 min
Delta R.T.: -0.001 min
Response: 112440085
Conc: 49.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.933 min
Delta R.T.: -0.001 min
Response: 180248207
Conc: 55.48 ng/ml




#16 4,4'-DDD

R.T.: 6.709 min
Delta R.T.: -0.002 min
Response: 97729958
Conc: 55.66 ng/ml



#16 4,4'-DDD


R.T.: 5.786 min
Delta R.T.: -0.002 min
Response: 163344583
Conc: 57.72 ng/ml



#17 4,4'-DDT


R.T.: 7.024 min
Delta R.T.: 0.000 min
Response: 96544968
Conc: 52.23 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.037 min
Delta R.T.: 0.000 min
Response: 164308313
Conc: 54.40 ng/ml




#18 Endrin aldehyde

R.T.: 6.924 min
Delta R.T.: -0.002 min
Response: 91177159
Conc: 51.38 ng/ml



#18 Endrin aldehyde


R.T.: 6.112 min
Delta R.T.: 0.000 min
Response: 144971365
Conc: 53.83 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.158 min
Delta R.T.: -0.001 min
Response: 104616518
Conc: 51.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.335 min
Delta R.T.: -0.002 min
Response: 171983119
Conc: 54.52 ng/ml




#20 Methoxychlor

R.T.: 7.500 min
Delta R.T.: 0.000 min
Response: 53535568
Conc: 53.55 ng/ml



#20 Methoxychlor


R.T.: 6.611 min
Delta R.T.: -0.001 min
Response: 87904365
Conc: 54.61 ng/ml



#21 Endrin ketone


R.T.: 7.643 min
Delta R.T.: -0.001 min
Response: 116927916
Conc: 52.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.840 min
Delta R.T.: -0.001 min
Response: 196755586
Conc: 54.05 ng/ml




#22 Mirex

R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 94046492
Conc: 50.33 ng/ml



#22 Mirex


R.T.: 7.020 min
Delta R.T.: -0.001 min
Response: 156801604
Conc: 51.30 ng/ml



#24 Chlordane-2


R.T.: 5.257 min
Delta R.T.: 0.026 min
Response: 151911125
Conc: 1381.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 4.365 min
Delta R.T.: 0.015 min
Response: 5474380
Conc: 39.44 ng/ml




#25 Chlordane-3

R.T.: 5.940 min
Delta R.T.: -0.001 min
Response: 130315187
Conc: 342.87 ng/ml



#25 Chlordane-3


R.T.: 4.979 min
Delta R.T.: 0.000 min
Response: 211556858
Conc: 498.53 ng/ml



#26 Chlordane-4


R.T.: 6.019 min
Delta R.T.: -0.004 min
Response: 130570041
Conc: 287.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 207265394
Conc: 503.88 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.872 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.933 min
Delta R.T.: -0.005 min
Response: 180248207
Conc: 1353.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.054 min
Delta R.T.: -0.002 min
Response: 95503809
Conc: 51.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.912 min
Delta R.T.: 0.000 min
Response: 158597626
Conc: 53.12 ng/ml