

Surrogate Summary

SDG No.: **P5380**

Client: **Weston Solutions**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093481.D	PIBLK-PL093481.D	Decachlorobiphenyl	1	20	22.0	110		30	135
		Tetrachloro-m-xylene	1	20	20.5	102		44	124
		Decachlorobiphenyl	2	20	21.4	107		30	135
		Tetrachloro-m-xylene	2	20	19.9	100		44	124
I.BLK-PL093540.D	PIBLK-PL093540.D	Decachlorobiphenyl	1	20	24.2	121		30	135
		Tetrachloro-m-xylene	1	20	23.3	116		44	124
		Decachlorobiphenyl	2	20	23.1	115		30	135
		Tetrachloro-m-xylene	2	20	22.3	112		44	124
PB165895BL	PB165895BL	Decachlorobiphenyl	1	20	22.0	110		30	135
		Tetrachloro-m-xylene	1	20	19.2	96		44	124
		Decachlorobiphenyl	2	20	21.0	105		30	135
		Tetrachloro-m-xylene	2	20	18.8	94		44	124
PB165895BS	PB165895BS	Decachlorobiphenyl	1	20	19.8	99		30	135
		Tetrachloro-m-xylene	1	20	18.0	90		44	124
		Decachlorobiphenyl	2	20	19.9	100		30	135
		Tetrachloro-m-xylene	2	20	17.1	86		44	124
PB165858TB	PB165858TB	Decachlorobiphenyl	1	20	22.3	112		30	135
		Tetrachloro-m-xylene	1	20	20.4	102		44	124
		Decachlorobiphenyl	2	20	23.0	115		30	135
		Tetrachloro-m-xylene	2	20	19.6	98		44	124
P5362-02MS	WC-SOIL-20241219MS	Decachlorobiphenyl	1	20	17.9	89		30	135
		Tetrachloro-m-xylene	1	20	19.1	95		44	124
		Decachlorobiphenyl	2	20	18.3	91		30	135
		Tetrachloro-m-xylene	2	20	19.3	97		44	124
P5362-02MSD	WC-SOIL-20241219MSD	Decachlorobiphenyl	1	20	17.6	88		30	135
		Tetrachloro-m-xylene	1	20	18.9	95		44	124
		Decachlorobiphenyl	2	20	18.1	90		30	135
		Tetrachloro-m-xylene	2	20	18.9	94		44	124
P5380-02	TAPIAL3-IDW-SOIL-122024-T1	Decachlorobiphenyl	1	20	15.8	79		30	135
		Tetrachloro-m-xylene	1	20	16.4	82		44	124
		Decachlorobiphenyl	2	20	16.1	81		30	135
		Tetrachloro-m-xylene	2	20	16.3	82		44	124
I.BLK-PL093553.D	PIBLK-PL093553.D	Decachlorobiphenyl	1	20	23.7	119		30	135
		Tetrachloro-m-xylene	1	20	22.4	112		44	124
		Decachlorobiphenyl	2	20	24.1	120		30	135
		Tetrachloro-m-xylene	2	20	21.6	108		44	124