



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5380

Client: Weston Solutions

Analytical Method: 8081B Datafile : PL093547.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165895BS	gamma-BHC (Lindane)	0.5	0.43	ug/L	86				59	134	
	Heptachlor	0.5	0.45	ug/L	90				54	130	
	Heptachlor epoxide	0.5	0.43	ug/L	87				61	133	
	Endrin	0.5	0.44	ug/L	88				60	138	
	Methoxychlor	0.5	0.44	ug/L	89				54	145	