

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF122624

Review By	yogesh	Review On	12/27/2024 4:07:20 AM
Supervise By	mohammad	Supervise On	12/27/2024 4:13:24 AM
SubDirectory	BF122624	HP Acquire Method	BNA_F
		HP Processing Method	bf122624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140969.D	26 Dec 2024 15:57		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140970.D	26 Dec 2024 16:23		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140971.D	26 Dec 2024 16:49	Compound #9,54,56,65 removed from 5ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF140972.D	26 Dec 2024 17:15		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF140973.D	26 Dec 2024 17:41	Compound #9,26,32,48,51,53,57,62,65 Kept on LR and Compound #54 Kept on QR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140974.D	26 Dec 2024 18:06	The Calibration is Good For 8270 DOD Except Com#56 and good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF140975.D	26 Dec 2024 18:33		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF140976.D	26 Dec 2024 18:59		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF140977.D	26 Dec 2024 19:25	Compound #9,69 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBF122624	BF140978.D	26 Dec 2024 19:52	ICV Fail for Com#56 for DOD	RC/JU	Ok,M
11	PB165543BL	PB165543BL	BF140979.D	26 Dec 2024 20:18		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF123024

Review By	Jagrut	Review On	12/31/2024 4:34:29 PM
Supervise By	mohammad	Supervise On	1/1/2025 7:15:14 AM
SubDirectory	BF123024	HP Acquire Method	BNA_F
		HP Processing Method	bf122624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141018.D	30 Dec 2024 08:58		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141019.D	30 Dec 2024 09:24		RC/JU	Ok
3	PB165894BL	PB165894BL	BF141020.D	30 Dec 2024 09:50		RC/JU	Ok
4	PB165894BS	PB165894BS	BF141021.D	30 Dec 2024 10:17		RC/JU	Ok,M
5	PB165858TB	PB165858TB	BF141022.D	30 Dec 2024 10:43		RC/JU	Ok
6	PB165875BS	PB165875BS	BF141023.D	30 Dec 2024 11:10		RC/JU	Ok,M
7	PB165790BL	PB165790BL	BF141024.D	30 Dec 2024 11:36		RC/JU	Ok
8	PB165875BSD	PB165875BSD	BF141025.D	30 Dec 2024 12:03		RC/JU	Ok,M
9	PB165875BL	PB165875BL	BF141026.D	30 Dec 2024 12:30		RC/JU	Ok
10	PB165872BS	PB165872BS	BF141027.D	30 Dec 2024 12:56		RC/JU	Ok,M
11	P5380-02	TAPIAL3-IDW-SOIL-12	BF141028.D	30 Dec 2024 13:29		RC/JU	Ok
12	SSTDCCC040	SSTDCCC040EC	BF141029.D	30 Dec 2024 14:01		RC/JU	Ok
13	DFTPP	DFTPP	BF141030.D	30 Dec 2024 14:54		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF141031.D	30 Dec 2024 15:20		RC/JU	Ok
15	PB165900BL	PB165900BL	BF141032.D	30 Dec 2024 15:47		RC/JU	Ok
16	P5362-02	WC-SOIL-20241219	BF141033.D	30 Dec 2024 16:20		RC/JU	Ok
17	P5362-02MS	WC-SOIL-20241219MS	BF141034.D	30 Dec 2024 16:47	END CCC Missing	RC/JU	Not Ok
18	P5362-02MSD	WC-SOIL-20241219MS	BF141035.D	30 Dec 2024 17:13	END CCC Missing	RC/JU	Not Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF123024

Review By	Jagrut	Review On	12/31/2024 4:34:29 PM		
Supervise By	mohammad	Supervise On	1/1/2025 7:15:14 AM		
SubDirectory	BF123024	HP Acquire Method	BNA_F	HP Processing Method	bf122624
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673				
CCC	SP6676				
Internal Standard/PEM	S12646,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P5386-02	MOO-24-00398	BF141036.D	30 Dec 2024 17:39		RC/JU	Ok
20	P5386-04	MOO-24-00395-96	BF141037.D	30 Dec 2024 18:06		RC/JU	Ok
21	P5386-05	MOO-24-00397	BF141038.D	30 Dec 2024 18:32		RC/JU	Ok
22	P5386-01	MOO-24-00398	BF141039.D	30 Dec 2024 18:58	Internal Standard Fail	RC/JU	Ok
23	P5386-03	MOO-24-00395-96	BF141040.D	30 Dec 2024 19:25	Internal Standard Fail	RC/JU	Ok
24	P5391-01	NB-07-12272024	BF141041.D	30 Dec 2024 19:51		RC/JU	Ok,M
25	P5391-01MS	NB-07-12272024MS	BF141042.D	30 Dec 2024 20:18		RC/JU	Ok,M
26	P5391-01MSD	NB-07-12272024MSD	BF141043.D	30 Dec 2024 20:44		RC/JU	Ok,M
27	P5390-01	EO-01-12272024	BF141044.D	30 Dec 2024 21:11		RC/JU	Ok,M
28	P5318-01	AU-06-122024	BF141045.D	30 Dec 2024 21:37	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF123124

Review By	Jagrut	Review On	12/31/2024 4:33:25 PM
Supervise By	mohammad	Supervise On	1/3/2025 12:31:08 AM
SubDirectory	BF123124	HP Acquire Method	BNA_F
		HP Processing Method	bf122624

STD. NAME	STD REF.#
Tune/Reschk	SP6701
Initial Calibration Stds	SP6680,SP6679,SP6678,SP6677,SP6676,SP6675,SP6674,SP6673
CCC	SP6676
Internal Standard/PEM	S12646,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141046.D	31 Dec 2024 09:24		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141047.D	31 Dec 2024 09:50		RC/JU	Ok
3	PB165790BL	PB165790BL	BF141048.D	31 Dec 2024 10:17		RC/JU	Ok
4	PB165900BS	PB165900BS	BF141049.D	31 Dec 2024 10:44	Surrogate Fail, Recovery Fail High For Many Compound	RC/JU	Not Ok
5	P5362-02MS	WC-SOIL-20241219MS	BF141050.D	31 Dec 2024 11:23		RC/JU	Ok,M
6	P5362-02MSD	WC-SOIL-20241219MS	BF141051.D	31 Dec 2024 11:49		RC/JU	Ok,M
7	P5398-03	352	BF141052.D	31 Dec 2024 12:16		RC/JU	Ok
8	P5398-01	341-349	BF141053.D	31 Dec 2024 12:42		RC/JU	Ok,M
9	P5398-01MS	341-349MS	BF141054.D	31 Dec 2024 13:09		RC/JU	Ok,M
10	P5398-01MSD	341-349MSD	BF141055.D	31 Dec 2024 13:35		RC/JU	Ok,M
11	SSTDCCC040	SSTDCCC040EC	BF141056.D	31 Dec 2024 14:18		RC/JU	Ok,M

M : Manual Integration