

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5380

Client: Weston Solutions

Analytical Method: 8270E DataFile: BF141021.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165894BS	Pyridine	50	45.5	ug/L	91				10	153	
	1,4-Dichlorobenzene	50	47.7	ug/L	95				29	112	
	2-Methylphenol	50	50.1	ug/L	100				30	117	
	3+4-Methylphenols	50	48.9	ug/L	98				29	110	
	Hexachloroethane	50	49.2	ug/L	98				21	115	
	Nitrobenzene	50	51.7	ug/L	103				45	121	
	Hexachlorobutadiene	50	48.2	ug/L	96				22	124	
	2,4,6-Trichlorophenol	50	52.9	ug/L	106				50	125	
	2,4,5-Trichlorophenol	50	53.4	ug/L	107				53	123	
	2,4-Dinitrotoluene	50	55.4	ug/L	111				57	128	
	Hexachlorobenzene	50	54.3	ug/L	109				53	125	
	Pentachlorophenol	100	110	ug/L	110				35	138	