

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: P5380 SAS No.: P5380 SDG No.: P5380

Instrument ID: MSVOA_N Calibration Date/Time: 12/27/2024 19:54

Lab File ID: VN085337.D Init. Calib. Date(s): 12/27/2024 12/27/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:26 12:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.804	0.723		-10.07	50
1,1-Dichloroethene	0.556	0.532		-4.32	50
2-Butanone	0.364	0.383		5.22	50
Carbon Tetrachloride	0.498	0.476		-4.42	50
Chloroform	1.181	1.084		-8.21	50
Benzene	1.381	1.318		-4.56	50
1,2-Dichloroethane	0.495	0.465		-6.06	50
Trichloroethene	0.327	0.295		-9.79	50
Tetrachloroethene	0.317	0.281		-11.36	50
Chlorobenzene	1.080	1.022	0.3	-5.37	50
1,2-Dichloroethane-d4	0.713	0.722		1.26	50
Dibromofluoromethane	0.312	0.320		2.56	50
Toluene-d8	1.145	1.166		1.83	50
4-Bromofluorobenzene	0.380	0.390		2.63	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.