

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: **P5380**

Client: **Weston Solutions**

Analytical Method: **SW8260D**

Datafile : VN085330.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1227WBS01	Vinyl chloride	20	16.8	ug/L	84			58	137	
	1,1-Dichloroethene	20	17.4	ug/L	87			71	131	
	2-Butanone	100	96.2	ug/L	96			56	143	
	Carbon Tetrachloride	20	17.6	ug/L	88			72	136	
	Chloroform	20	17.4	ug/L	87			79	124	
	Benzene	20	17.5	ug/L	88			79	120	
	1,2-Dichloroethane	20	17.3	ug/L	86			73	128	
	Trichloroethene	20	16.8	ug/L	84			79	123	
	Tetrachloroethene	20	18.0	ug/L	90			74	129	
	Chlorobenzene	20	17.7	ug/L	89			82	118	