

SOP ID : MSM4500-NH3 B,G-Ammonia-17

SDG No : N/A

Start Digest Date: 12/26/2024 Time : 08:30 Temp : 150 °C

Matrix : SOIL

End Digest Date: 12/26/2024 Time : 09:30 Temp : 160 °C

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: RM

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: 12

Standard Name	MLS USED	STD REF. # FROM LOG
LCSS	1.0ML	WP111092
MS/MSD SPIKE SOL.	1.0ML	WP111091
PBS003	50.0ML	W3112
RL CHECK	0.1ML	WP111091
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP108708
NAOH 6N	0.5-2.0ML	WP108660
H2SO4 0.04N	5.0ML	WP110335
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT
WP108814

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/26/2024 09:45	RM (w/c)	RM (w/c)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
P5382-01DUP	COMP-1DUP	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
P5382-01MS	COMP-1MS	1.02	50	N/A	N/A	N/A	N/A	N/A	N/A
P5382-01MSD	COMP-1MSD	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
P5382-01	COMP-1	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
P5382-02	COMP-2	1.04	50	N/A	N/A	N/A	N/A	N/A	N/A
P5382-03	COMP-3	1.01	50	N/A	N/A	N/A	N/A	N/A	N/A
PB165837BL	PBS837	1.00	50	N/A	N/A	N/A	N/A	N/A	N/A
PB165837BS	LCS837	1.00	50	N/A	N/A	N/A	N/A	N/A	N/A