

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: P5382

Project: Comegys School

RunNo.: LB134081

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.3	10	103	90-110	12/18/2024
Chloride	mg/L	3.1	3	103	90-110	12/18/2024
Fluoride	mg/L	2	2	100	90-110	12/18/2024
Nitrite	mg/L	3.1	3	103	90-110	12/18/2024
Nitrate	mg/L	2.6	2.5	104	90-110	12/18/2024
Sulfate	mg/L	15.3	15	102	90-110	12/18/2024
Orthophosphate as P	mg/L	5.1	5	102	90-110	12/18/2024
Sample ID: CCV1						
Bromide	mg/L	10.4	10	104	90-110	12/26/2024
Chloride	mg/L	3.1	3	103	90-110	12/26/2024
Fluoride	mg/L	2.1	2	105	90-110	12/26/2024
Nitrite	mg/L	3.1	3	103	90-110	12/26/2024
Nitrate	mg/L	2.6	2.5	104	90-110	12/26/2024
Sulfate	mg/L	15.7	15	105	90-110	12/26/2024
Orthophosphate as P	mg/L	4.9	5	98	90-110	12/26/2024
Sample ID: CCV2						
Bromide	mg/L	10.4	10	104	90-110	12/26/2024
Chloride	mg/L	3.1	3	103	90-110	12/26/2024
Fluoride	mg/L	2.1	2	105	90-110	12/26/2024
Nitrite	mg/L	3.1	3	103	90-110	12/26/2024
Nitrate	mg/L	2.6	2.5	104	90-110	12/26/2024
Sulfate	mg/L	15.5	15	103	90-110	12/26/2024
Orthophosphate as P	mg/L	5.3	5	106	90-110	12/26/2024

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: P5382

Project: Comegys School

RunNo.: LB134085

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	12/26/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	12/26/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	12/26/2024

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: P5382

Project: Comegys School

RunNo.: LB134087

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024
Sample ID: CCV1 Ammonia as N	mg/L	0.98	1	98	90-110	12/26/2024
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024