

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134086

Review By	mohan	Review On	12/26/2024 10:16:20 PM
Supervise By	jaswal	Supervise On	12/27/2024 8:27:30 PM

STD. NAME	STD REF.#
ICAL Standard	MP83766,MP83767,MP83768,MP83769,MP83770,MP83771
ICV Standard	MP83772
CCV Standard	MP83774
ICSA Standard	
CRI Standard	MP83776
LCS Standard	
Chk Standard	MP83773,MP83775,MP83777,MP83779

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/26/24 12:48		mohan	OK
2	S0.2	S0.2	CAL2	12/26/24 12:50		mohan	OK
3	S2.5	S2.5	CAL3	12/26/24 12:52		mohan	OK
4	S5	S5	CAL4	12/26/24 12:55		mohan	OK
5	S7.5	S7.5	CAL5	12/26/24 12:57		mohan	OK
6	S10	S10	CAL6	12/26/24 13:05		mohan	OK
7	ICV31	ICV31	ICV	12/26/24 13:10		mohan	OK
8	ICB31	ICB31	ICB	12/26/24 13:13		mohan	OK
9	CCV67	CCV67	CCV	12/26/24 13:15		mohan	OK
10	CCB67	CCB67	CCB	12/26/24 13:17		mohan	OK
11	CRA	CRA	CRDL	12/26/24 13:19		mohan	OK
12	HighStd	HighStd	HIGH STD	12/26/24 13:22		mohan	OK
13	ChkStd	ChkStd	SAM	12/26/24 13:24		mohan	OK
14	PB165862BL	PB165862BL	MB	12/26/24 13:26		mohan	OK
15	PB165862BS	PB165862BS	LCS	12/26/24 13:29		mohan	OK
16	P5318-01	AU-06-122024	SAM	12/26/24 13:31		mohan	OK
17	P5382-01	COMP-1	SAM	12/26/24 13:40		mohan	OK
18	P5382-01DUP	COMP-1DUP	DUP	12/26/24 13:42		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134086

Review By	mohan	Review On	12/26/2024 10:16:20 PM
Supervise By	jaswal	Supervise On	12/27/2024 8:27:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83766,MP83767,MP83768,MP83769,MP83770,MP83771		
ICV Standard	MP83772		
CCV Standard	MP83774		
ICSA Standard			
CRI Standard	MP83776		
LCS Standard			
Chk Standard	MP83773,MP83775,MP83777,MP83779		

19	P5382-01MS	COMP-1MS	MS	12/26/24 13:47		mohan	OK
20	P5382-01MSD	COMP-1MSD	MSD	12/26/24 13:49		mohan	OK
21	CCV68	CCV68	CCV	12/26/24 13:52		mohan	OK
22	CCB68	CCB68	CCB	12/26/24 13:54		mohan	OK
23	P5382-02	COMP-2	SAM	12/26/24 13:56		mohan	OK
24	P5382-03	COMP-3	SAM	12/26/24 13:59		mohan	OK
25	P5383-01	OK-02-12232024	SAM	12/26/24 14:01		mohan	OK
26	P5382-01L	COMP-1L	SD	12/26/24 14:06		mohan	OK
27	P5382-01A	COMP-1A	PS	12/26/24 14:09		mohan	OK
28	CCV69	CCV69	CCV	12/26/24 14:12		mohan	OK
29	CCB69	CCB69	CCB	12/26/24 14:14		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/30/24 13:54		Kareem	OK
2	S1	S1	CAL2	12/30/24 13:58		Kareem	OK
3	S2	S2	CAL3	12/30/24 14:03		Kareem	OK
4	S3	S3	CAL4	12/30/24 14:07		Kareem	OK
5	S4	S4	CAL5	12/30/24 14:11		Kareem	OK
6	S5	S5	CAL6	12/30/24 14:15		Kareem	OK
7	ICV01	ICV01	ICV	12/30/24 14:20		Kareem	OK
8	LLICV01	LLICV01	LLICV	12/30/24 14:24		Kareem	OK
9	ICB01	ICB01	ICB	12/30/24 14:28		Kareem	OK
10	CRI01	CRI01	CRDL	12/30/24 14:32		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/30/24 14:37		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/30/24 14:45		Kareem	OK
13	CCV01	CCV01	CCV	12/30/24 14:49		Kareem	OK
14	CCB01	CCB01	CCB	12/30/24 14:53		Kareem	OK
15	P5342-08	CHRT26634	SAM	12/30/24 14:58		Kareem	OK
16	P5342-09	HT2651	SAM	12/30/24 15:02		Kareem	OK
17	P5382-01	COMP-1	SAM	12/30/24 15:06		Kareem	OK
18	P5382-02	COMP-2	SAM	12/30/24 15:11		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

19	P5382-03	COMP-3	SAM	12/30/24 15:15		Kareem	OK
20	CCV02	CCV02	CCV	12/30/24 15:19		Kareem	OK
21	CCB02	CCB02	CCB	12/30/24 15:23		Kareem	OK
22	P5362-01	WC-SOIL-20241219	SAM	12/30/24 15:28		Kareem	OK
23	P5362-01DUP	WC-SOIL-20241219D	DUP	12/30/24 15:35		Kareem	OK
24	P5362-01L	WC-SOIL-20241219L	SD	12/30/24 15:39		Kareem	OK
25	P5362-01MS	WC-SOIL-20241219M	MS	12/30/24 15:43	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
26	P5362-01MSD	WC-SOIL-20241219M	MSD	12/30/24 15:47	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
27	P5362-01A	WC-SOIL-20241219A	PS	12/30/24 15:51	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
28	P5386-01	MOO-24-00398	SAM	12/30/24 15:55		Kareem	OK
29	P5386-03	MOO-24-00395-96	SAM	12/30/24 15:59		Kareem	OK
30	P5318-01	AU-06-122024	SAM	12/30/24 16:03		Kareem	OK
31	P5387-01	TR-05-122624	SAM	12/30/24 16:08		Kareem	OK
32	CCV03	CCV03	CCV	12/30/24 16:12		Kareem	OK
33	CCB03	CCB03	CCB	12/30/24 16:16		Kareem	OK
34	PB165863BL	PB165863BL	MB	12/30/24 16:20		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

35	PB165863BS	PB165863BS	LCS	12/30/24 16:25	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
36	P5362-02	WC-SOIL-20241219	SAM	12/30/24 16:29		Kareem	OK
37	P5380-02	TAPIAL3-IDW-SOIL-1	SAM	12/30/24 16:33		Kareem	OK
38	P5386-02	MOO-24-00398	SAM	12/30/24 16:38		Kareem	OK
39	P5386-04	MOO-24-00395-96	SAM	12/30/24 16:42		Kareem	OK
40	P5386-04DUP	MOO-24-00395-96DUP	DUP	12/30/24 16:47		Kareem	OK
41	P5386-04L	MOO-24-00395-96L	SD	12/30/24 16:51		Kareem	OK
42	P5386-04MS	MOO-24-00395-96MS	MS	12/30/24 16:55	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
43	P5386-04MSD	MOO-24-00395-96MSD	MSD	12/30/24 16:59	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
44	CCV04	CCV04	CCV	12/30/24 17:04		Kareem	OK
45	CCB04	CCB04	CCB	12/30/24 17:08		Kareem	OK
46	P5386-04A	MOO-24-00395-96A	PS	12/30/24 17:12	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
47	PB165858TB	PB165858TB	MB	12/30/24 17:16	LCS fail for Ba	Kareem	Not Ok
48	PB165887BL	PB165887BL	MB	12/30/24 17:21	LCS fail for Ba	Kareem	Not Ok
49	PB165887BS	PB165887BS	LCS	12/30/24 17:25	LCS fail for Ba	Kareem	Not Ok
50	P5356-01	STAND-PIPE	SAM	12/30/24 17:29		Kareem	OK
51	P5356-01DUP	STAND-PIPEDUP	DUP	12/30/24 17:33		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

52	P5356-01L	STAND-PIPEL	SD	12/30/24 17:38		Kareem	OK
53	P5356-01MS	STAND-PIPEMS	MS	12/30/24 17:42	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
54	P5356-01MSD	STAND-PIPEMSD	MSD	12/30/24 17:46	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
55	P5356-01A	STAND-PIPEA	PS	12/30/24 17:50	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
56	CCV05	CCV05	CCV	12/30/24 17:54		Kareem	OK
57	CCB05	CCB05	CCB	12/30/24 17:58		Kareem	OK
58	P5377-02DL	GAS-TRE-1114DL	SAM	12/30/24 18:02	Confirm wt/vol	Kareem	OK
59	P5386-05	MOO-24-00397	SAM	12/30/24 18:07		Kareem	OK
60	PB165879BL	PB165879BL	MB	12/30/24 18:11		Kareem	OK
61	PB165879BS	PB165879BS	LCS	12/30/24 18:15	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
62	P5390-01	EO-01-12272024	SAM	12/30/24 18:19		Kareem	OK
63	P5391-01	NB-07-12272024	SAM	12/30/24 18:23		Kareem	OK
64	P5391-01DUP	NB-07-12272024DUP	DUP	12/30/24 18:28		Kareem	OK
65	P5391-01L	NB-07-12272024L	SD	12/30/24 18:32		Kareem	OK
66	P5391-01MS	NB-07-12272024MS	MS	12/30/24 18:36	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134129

Review By	kareem	Review On	12/31/2024 7:20:20 PM
Supervise By	mohan	Supervise On	12/31/2024 11:26:17 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

67	P5391-01MSD	NB-07-12272024MSD	MSD	12/30/24 18:40	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
68	CCV06	CCV06	CCV	12/30/24 18:44		Kareem	OK
69	CCB06	CCB06	CCB	12/30/24 18:48		Kareem	OK
70	P5391-01A	NB-07-12272024A	PS	12/30/24 18:53	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
71	PB165917BL	PB165917BL	MB	12/30/24 18:57		Kareem	OK
72	PB165917BS	PB165917BS	LCS	12/30/24 19:01	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
73	LR1	LR1	HIGH STD	12/30/24 19:14		Kareem	OK
74	LR2	LR2	HIGH STD	12/30/24 19:19		Kareem	OK
75	CCV07	CCV07	CCV	12/30/24 19:24		Kareem	OK
76	CCB07	CCB07	CCB	12/30/24 19:31		Kareem	OK