



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Kleinfelder **SDG No.:** P5382
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5382 **SAS No.:** P5382
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.16	0.2	82	40 - 160	CV	12/26/2024	13:19	LB134086
CRI01	Aluminum	93.7	100	94	40 - 160	P	12/30/2024	14:32	LB134129
	Antimony	46.6	50.0	93	40 - 160	P	12/30/2024	14:32	LB134129
	Arsenic	21.1	20.0	105	40 - 160	P	12/30/2024	14:32	LB134129
	Barium	91.3	100	91	40 - 160	P	12/30/2024	14:32	LB134129
	Beryllium	5.68	6.0	95	40 - 160	P	12/30/2024	14:32	LB134129
	Boron	97.7	100	98	40 - 160	P	12/30/2024	14:32	LB134129
	Cadmium	6.07	6.0	101	40 - 160	P	12/30/2024	14:32	LB134129
	Chromium	9.09	10.0	91	40 - 160	P	12/30/2024	14:32	LB134129
	Cobalt	29.0	30.0	96	40 - 160	P	12/30/2024	14:32	LB134129
	Copper	20.6	20.0	103	40 - 160	P	12/30/2024	14:32	LB134129
	Iron	95.3	100	95	40 - 160	P	12/30/2024	14:32	LB134129
	Lead	11.3	12.0	94	40 - 160	P	12/30/2024	14:32	LB134129
	Manganese	19.2	20.0	96	40 - 160	P	12/30/2024	14:32	LB134129
	Molybdenum	197	200	99	40 - 160	P	12/30/2024	14:32	LB134129
	Nickel	38.7	40.0	97	40 - 160	P	12/30/2024	14:32	LB134129
	Selenium	17.9	20.0	89	40 - 160	P	12/30/2024	14:32	LB134129
	Silver	10.4	10.0	104	40 - 160	P	12/30/2024	14:32	LB134129
	Thallium	37.1	40.0	93	40 - 160	P	12/30/2024	14:32	LB134129
	Vanadium	37.7	40.0	94	40 - 160	P	12/30/2024	14:32	LB134129
	Zinc	41.1	40.0	103	40 - 160	P	12/30/2024	14:32	LB134129