

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder **SDG No.:** P5382
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5382 **SAS No.:** P5382

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165862BS Mercury	mg/Kg	0.25	0.22		87	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder **SDG No.:** P5382
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5382 **SAS No.:** P5382

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165863BS							
Aluminum	mg/Kg	83.3	74.2		89	80 - 120	P
Antimony	mg/Kg	33.3	31.7		95	80 - 120	P
Arsenic	mg/Kg	33.3	32.8		98	80 - 120	P
Barium	mg/Kg	8.3	7.41		89	80 - 120	P
Beryllium	mg/Kg	8.3	8.17		98	80 - 120	P
Boron	mg/Kg	12.5	11.8		94	80 - 120	P
Cadmium	mg/Kg	8.3	8.30		100	80 - 120	P
Chromium	mg/Kg	16.7	16.7		100	80 - 120	P
Cobalt	mg/Kg	8.3	8.11		98	80 - 120	P
Copper	mg/Kg	12.5	12.3		98	80 - 120	P
Iron	mg/Kg	130	121		93	80 - 120	P
Lead	mg/Kg	41.7	40.1		96	80 - 120	P
Manganese	mg/Kg	8.3	7.82		94	80 - 120	P
Molybdenum	mg/Kg	16.7	16.0		96	80 - 120	P
Nickel	mg/Kg	20.8	20.5		99	80 - 120	P
Selenium	mg/Kg	83.3	81.1		97	80 - 120	P
Silver	mg/Kg	3.1	3.13		101	80 - 120	P
Thallium	mg/Kg	83.3	78.6		94	80 - 120	P
Vanadium	mg/Kg	12.5	11.6		93	80 - 120	P
Zinc	mg/Kg	8.3	8.16		98	80 - 120	P