



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB165843BS

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

Lab Sample ID: PB165843BS Date(s) Analyzed: 12/26/2024 12/26/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO108764.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.798	4.748	4.848	157	165	4.33
COLUMN 1	2	4.818	4.768	4.868	160		
	3	4.874	4.824	4.924	173		
	4	4.996	4.946	5.046	171		
	5	5.253	5.203	5.303	163		
	1	4.786	4.736	4.836	156		
COLUMN 2	2	4.806	4.756	4.856	161	158	
	3	4.981	4.931	5.031	159		
	4	5.023	4.973	5.073	158		
	5	5.237	5.187	5.287	157		
Aroclor-1260	1	6.296	6.246	6.346	173	159	4.31
COLUMN 1	2	6.484	6.434	6.534	173		
	3	6.853	6.803	6.903	148		
	4	7.114	7.064	7.164	150		
	5	7.354	7.304	7.404	151		
	1	6.27	6.22	6.32	172		
COLUMN 2	2	6.456	6.406	6.506	172	166	
	3	6.61	6.56	6.66	171		
	4	7.082	7.032	7.132	156		
	5	7.321	7.271	7.371	158		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

AU-06-122024MS

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

Lab Sample ID: P5318-01MS Date(s) Analyzed: 12/26/2024 12/26/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO108767.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.799	4.749	4.849	140	128	7.52
COLUMN 1	2	4.819	4.769	4.869	129		
	3	4.874	4.824	4.924	134		
	4	4.995	4.945	5.045	131		
	5	5.253	5.203	5.303	109		
	1	4.786	4.736	4.836	139		
COLUMN 2	2	4.805	4.755	4.855	139	138	
	3	4.982	4.932	5.032	146		
	4	5.023	4.973	5.073	145		
	5	5.236	5.186	5.286	119		
Aroclor-1254	1	5.61	5.56	5.66	62.9	63.0	6.15
COLUMN 1	2	5.758	5.708	5.808	61.1		
	3	6.18	6.13	6.23	47.4		
	4	6.394	6.344	6.444	18.9		
	5	6.813	6.763	6.863	125		
	1	5.589	5.539	5.639	64.1		
COLUMN 2	2	5.735	5.685	5.785	60.2	67.0	
	3	6.155	6.105	6.205	43.5		
	4	6.367	6.317	6.417	22.0		
	5	6.784	6.734	6.834	145		



Aroclor-1260	1	6.296	6.246	6.346	121		
	2	6.485	6.435	6.535	115		
	3	6.854	6.804	6.904	87.1		
	4	7.113	7.063	7.163	90.3		
	5	7.355	7.305	7.405	89.9		
COLUMN 1						101	20.44
	1	6.27	6.22	6.32	137		
	2	6.457	6.407	6.507	134		
	3	6.61	6.56	6.66	132		
	4	7.081	7.031	7.131	107		
5	7.321	7.271	7.371	107			
COLUMN 2						124	

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

AU-06-122024MSD

Contract: POWE02

Lab Code: CHEM

Case No.: P5382

SAS No.: P5382

SDG NO.: P5382

Lab Sample ID: P5318-01MSD

Date(s) Analyzed: 12/26/2024

12/26/2024

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO108768.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD	
			FROM	TO				
Aroclor-1016	1	4.798	4.748	4.848	137	131	2.26	
	2	4.817	4.767	4.867	132			
	3	4.873	4.823	4.923	138			
	4	4.995	4.945	5.045	133			
	5	5.253	5.203	5.303	114			
COLUMN 1								
	1	4.785	4.735	4.835	131	134		
	2	4.805	4.755	4.855	140			
	3	4.981	4.931	5.031	138			
	4	5.022	4.972	5.072	139			
5	5.236	5.186	5.286	124				
COLUMN 2								
	1	5.609	5.559	5.659	59.0	62.3	19.94	
	2	5.757	5.707	5.807	56.7			
	3	6.179	6.129	6.229	46.4			
	4	6.393	6.343	6.443	18.3			
5	6.813	6.763	6.863	131				
COLUMN 1								
	1	5.589	5.539	5.639	65.5	76.1		
	2	5.735	5.685	5.785	63.2			
	3	6.155	6.105	6.205	50.9			
	4	6.397	6.347	6.447	48.0			
5	6.783	6.733	6.833	153				
COLUMN 2								



Aroclor-1260	1	6.295	6.245	6.345	121		
	2	6.484	6.434	6.534	113		
	3	6.853	6.803	6.903	93.1		
	4	7.113	7.063	7.163	98.5		
	5	7.354	7.304	7.404	94.9		
COLUMN 1						104	19.91
	1	6.27	6.22	6.32	138		
	2	6.456	6.406	6.506	134		
	3	6.61	6.56	6.66	131		
	4	7.081	7.031	7.131	118		
5	7.321	7.271	7.371	114			
COLUMN 2						127	