

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

Continuing Calib Date: 12/26/2024 Initial Calibration Date(s): 12/23/2024 12/23/2024

Continuing Calib Time: 14:49 Initial Calibration Time(s): 13:15 14:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Dieldrin	6.35	6.35	6.25	6.45	0.01
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
4,4'-DDT	7.02	7.03	6.93	7.13	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

Continuing Calib Date: 12/26/2024 Initial Calibration Date(s): 12/23/2024 12/23/2024

Continuing Calib Time: 14:49 Initial Calibration Time(s): 13:15 14:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
Aldrin	4.23	4.23	4.13	4.33	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
4,4'-DDD	5.79	5.79	5.69	5.89	0.00
4,4'-DDT	6.04	6.04	5.94	6.14	0.00