

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/26/2024

Lab Sample No.: PSTDCCC050 Data File : PL093516.D Time Analyzed: 12:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.719	6.611	6.811	52.470	50.000	4.9
4,4'-DDE	6.202	6.094	6.294	53.250	50.000	6.5
4,4'-DDT	7.033	6.925	7.125	55.070	50.000	10.1
Aldrin	5.267	5.159	5.359	52.220	50.000	4.4
Decachlorobiphenyl	9.065	8.956	9.156	54.640	50.000	9.3
Dieldrin	6.354	6.246	6.446	53.020	50.000	6.0
Tetrachloro-m-xylene	3.548	3.442	3.642	50.880	50.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL01 Date Analyzed: 12/26/2024

Lab Sample No.: PSTDCCC050 Data File : PL093516.D Time Analyzed: 12:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.789	5.688	5.888	56.140	50.000	12.3
4,4'-DDE	5.235	5.133	5.333	53.970	50.000	7.9
4,4'-DDT	6.040	5.937	6.137	58.000	50.000	16.0
Aldrin	4.229	4.128	4.328	52.000	50.000	4.0
Decachlorobiphenyl	7.915	7.812	8.012	57.100	50.000	14.2
Dieldrin	5.366	5.264	5.464	51.400	50.000	2.8
Tetrachloro-m-xylene	2.777	2.677	2.877	50.790	50.000	1.6