

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/26/2024

Lab Sample No.: PSTDCCC050 Data File : PL093526.D Time Analyzed: 14:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.711	6.611	6.811	54.250	50.000	8.5
4,4'-DDE	6.194	6.094	6.294	53.270	50.000	6.5
4,4'-DDT	7.024	6.925	7.125	54.770	50.000	9.5
Aldrin	5.259	5.159	5.359	52.070	50.000	4.1
Decachlorobiphenyl	9.056	8.956	9.156	52.190	50.000	4.4
Dieldrin	6.345	6.246	6.446	52.070	50.000	4.1
Tetrachloro-m-xylene	3.540	3.442	3.642	51.770	50.000	3.5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/23/2024 12/23/2024

Client Sample No.: CCAL02 Date Analyzed: 12/26/2024

Lab Sample No.: PSTDCCC050 Data File : PL093526.D Time Analyzed: 14:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.787	5.688	5.888	56.180	50.000	12.4
4,4'-DDE	5.232	5.133	5.333	54.560	50.000	9.1
4,4'-DDT	6.037	5.937	6.137	56.650	50.000	13.3
Aldrin	4.227	4.128	4.328	54.060	50.000	8.1
Decachlorobiphenyl	7.912	7.812	8.012	54.840	50.000	9.7
Dieldrin	5.363	5.264	5.464	53.810	50.000	7.6
Tetrachloro-m-xylene	2.775	2.677	2.877	52.820	50.000	5.6